

## $SAMI^2 CO_2$ Calibration Certificate

June 14, 2023

Serial Number: C0090 RMA Number: 082302 Technician: BW

Reagent: btb\_0508

Table 1:  $SAMI^2 CO_2$  calibration parameters.

A	В	$\mathbf{C}$	$K_{434}$	$K_{620}$	$T_{avg}$	Range (ppm)
0.0374	0.6587	-1.5364	0.9201	1.3656	10.7242	202-1885

Table 2: Blank intensities.

Source	Reference	Signal
Dark	107	93
434  nm	3070	2820
$620~\mathrm{nm}$	1944	2600

## **Build Notes**

Pressure Test: 200 psi

**Modifications:** 

1

Standard Refurb

IR testing

ext cal range

