



SAMI^2 pH Validation Certificate

October 11, 2019

Serial Number: P0154 RMA Number: 082727

> Technician: B. Montgomery Reagent: PmCP_0569

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
18103.0	2296.0	97.8	40427.0

Validation Notes

Tris Batch: 070319 $\Delta T = 0.03$ °C

Tris Bottle: 3 N=9

Validation Temp: 25.02 °C Δ pH = 0.0044 ± 0.0007

Build Notes

Pressure Test: 200 psi - Passed

 ${\bf Modifications:}$

Standard refurbishment