



## SAMI<sup>2</sup> pH Validation Certificate

November 11, 2019

Serial Number: P0116 RMA Number: 082730 Technician: BW

Reagent: pmcp0583

Table 1:  $SAMI^2$  pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	$\mathrm{Eb}_{578}$
17372.0	2284.0	94.1	38676.0

## Validation Notes

Tris Batch: 011519  $\Delta T = 0.12\,^{\circ}\mathrm{C}$ 

Tris Bottle: 4 N=7

Validation Temp: 25.22 °C  $\Delta pH = -0.0026 \pm 0.0009$ 

## **Build Notes**

 $Pressure\ Test:\ 200\,psi-Passed$ 

 ${\bf Modifications:}$ 

Refurbishment with OOI 3-pin upgrade; PV cable, bulkhead, pressure plate and labor.

reagent was  ${\sim}90\%$  full