

## Certificate of Calibration Ocean Sonics Ltd.

## Calibration Number 3007

Unit Type
Serial Number
Build Date
Measured Date
Release Date
Sensitivity @ 26 Hz
Case Type
Element Number
F/W Version
Calibrated By
Work Order #


SB2-ETH
1249
16 March 2013
05 August 2013
05 August 2013
-168.49 dBV re μPa
Titanium
626
19b
T.Sandhu
W1033

Projector Model
Projector Serial
Projector output
Projector Mode
Distance
No. of Averages
Analog Reference
Ang. Ref. Model
Ang. Ref. Serial
Ref. Preamp
Ref. Preamp Serial
Ref. Preamp Gain
Digital Reference
Digital Ref. Serial

TH2-SER-4F
2206
131.5 dB re µPa
Wideband
0.5 Meter
2
Reson
TC4034-1
4211050
300415-01
78
30.0 dB
300422-01
87

Test Performed Facility
Test Result

Receive Sensitivity Calibration Calibration Tank #1 0 to 100 KHz, -170.5 ± 1.6 dB 0 to 200 KHz, -170.4 ± 4.9 dB



Frequency kHz	Sensitivity dBV re µPa
13.5	-169.3
27.1	-169.6
40.6	-172.6
54.1	-171.8
67.7	-169.2
81.2	-170.4
94.7	-169.1
108.3	-171.6
121.8	-174.5
135.3	-171.9
148.8	-170.0
162.4	-166.7
175.9	-164.6
189.4	-171.0
200	-173.9