



April 24, 2019

SAMI-pH Accuracy Certificate

Sunburst SAMI Serial Number: P0113

OOI Identification Number:

RSN P/N: 4830-58337

RSN AT: ATAPL58337-00005

Date: 11/19/13

pH Evaluates:

Ea₄₃₄ = 17533

Ea₅₇₈ = 101

Eb₄₃₄ = 2229

Eb₅₇₈ = 38502

Accuracy Test Temperature

T = 25 C

Toffset = 0 C

Certified Tris Batch #: Tris 102913_bag1

Tris Bottle/Bag #: Bag 1

Tris Correction: -0.0029

mCP lot: MKBC2604V

Date: 11/18/13

Pressure Test:

200psi – passed

Modifications:

- P/V lid with bubble trap
- Diaphragm retainer with ridge