

## $SAMI^2 CO_2$ Calibration Certificate

October 6, 2017

Serial Number: c78
RMA Number: 7160
Technician: B. Wasser
Reagent: btb0120

Table 1:  $SAMI^2 CO_2$  calibration parameters.

A	В	$\mathbf{C}$	$K_{434}$	$K_{620}$	$T_{avg}$	Range (ppm)
0.0750	0.4705	-1.3332	0.7989	0.9736	10.5150	203-1200

Table 2: Blank intensities.

Source	Reference	Signal
Dark	65	97
$434~\mathrm{nm}$	2786	2269
$620~\mathrm{nm}$	2379	2348

## **Build Notes**

Pressure Test: 200 psi

**Modifications:** 

XL reagent

Cold cal. 1200uATM 10C

