



## SAMI<sup>2</sup> pH Validation Certificate

November 25, 2019

Serial Number: P0117 RMA Number: 082730 Technician: BW

Reagent: pmcp0572

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
18103.0	2296.0	97.8	40427.0

## Validation Notes

Tris Batch: 101519  $\Delta T = 0.09$  °C

Tris Bottle: 5 N = 11

Validation Temp: 24.99 °C  $\Delta pH = -0.0035 \pm 0.0010$ 

## **Build Notes**

Pressure Test: 200 psi - Passed

 ${\bf Modifications:}$ 

Standard refurbishment

damaged pump and valve replaced