

Sunburst Sensors, LLC 1226 W Broadway Ave Missoula, MT 50802 +1 406 532 3247

## SAMI<sup>2</sup> CO<sub>2</sub> Calibration Certificate

April 15, 2016

Serial Number: ©124

OOI Numberi

UW-SC12-01-PHSEN-PCO2W-2004

Technician: MK

Reagent: 3-3-16

Table 1:  $SAMI^2 CO_2$  calibration parameters.

district and desired in	ACTIVITIES AND INCOME.	Control of the Contro	CONTRACTOR SO	CONTRACTOR AND A		・対応は、は、は、は、は、は、は、は、は、は、は、は、は、は、は、は、は、は、は、
A	В	Ç	$K_{434}$	$K_{020}$	$T_{avg}$	Range (ppm)
The second secon						
0.0357	0.7276	-1.8093	0.6246	1,5779	16.6406	200=1389

Table 2: Blank intensities,

Source	Reference	Signal
Dark	61	38
434 nm	3246	2030
$620~\mathrm{nm}$	1784	2756
CONTRACTOR OF STREET		CONTRACTOR CONTRACTOR

## **Bulld Notes**

Pressure Test: 200 psi

Modifications

