



April 13, 2015

## **OOI SAMI-CO<sub>2</sub> Calibration Certificate**

**Sunburst SAMI Serial Number: C0118**

**OOI Identification #: CGSN P/N: 1336-00012-000002**

**Date: 04/13/15**

**Calibration Range (ppm): 200-600**

### **CO<sub>2</sub> Calibration Coefficients:**

$a = 0.0767$

$b = 0.4333$

$c = -1.2901$

$K_{434} = 0.5760$

$K_{620} = 1.0527$

### **CO<sub>2</sub> Calibration Temperature**

$T = 14.6350$

### **Pressure Test:**

200 psi