



April 24, 2019

SAMI-pH Accuracy Certificate

Sunburst SAMI Serial Number: P0164

OOI Identification Number: 4830-58337

RSN P/N: ATAPL58337-00010

Date: 7/30/15

pH Evaluates:

$E_{a434} = 17533$

$E_{a578} = 101$

$E_{b434} = 2229$

$E_{b578} = 38502$

Accuracy Test Temperature

$T = 25\text{ C}$

$T_{\text{offset}} = 0.15\text{ C}$

Certified Tris Batch #: 061815

Tris Bottle/Bag #: Bag

Tris Correction: 0

mCP lot: MKBC2604V

Pressure Test:

200psi – passed

Technician: BW