



Sunburst Sensors, LLC
1226 W Broadway Ave
Missoula, MT 59802
+1 406 532 3247

SAMI² CO₂ Calibration Certificate

August 20, 2018

Serial Number: c91
RMA Number: 062913
Technician: B. Wasser
Reagent: btb0198

Table 1: SAMI² CO₂ calibration parameters.

A	B	C	K_{434}	K_{620}	T_{avg}	Range (ppm)
-0.0032	0.9523	-2.0758	1.0978	1.6270	5.8716	201-954

Table 2: Blank intensities.

Source	Reference	Signal
Dark	97	117
434 nm	3006	3311
620 nm	2063	3314

Build Notes

Pressure Test: 200 psi

Modifications:

60

REAGENT AT ~65% UPON ARRIVAL

