

## $SAMI^2 CO_2$ Calibration Certificate

May 16, 2023

Serial Number: C0075 RMA Number: 082343 Technician: BW

Reagent: btb\_0503

Table 1:  $SAMI^2 CO_2$  calibration parameters.

A	В	С	$K_{434}$	$K_{620}$	$T_{avg}$	Range (ppm)
0.0603	0.5538	-1.4517	1.3130	0.9752	10.6375	203-1187

Table 2: Blank intensities.

Source	Reference	Signal
Dark	46	32
434  nm	2462	3204
620  nm	3215	3125

## **Build Notes**

Pressure Test: 200 psi

**Modifications:** 

0

Standard Refurbishment

IR testing

New Cell and membrane holder

10 pct reagent

ext cal

