In Water Digital Linear PAR Sensor Calibration

Date:	April 28, 2015
Customer:	University of Washington
Project Number:	2015-7018
Model:	PAR LIN 600m
Serial Number:	0557

Calibration Coefficients

Im = 1.3589
a_0 = 2157207132.9 [count]
a_1 = 2.34482300090E-06 [µmol photons/m^2/s/count]

Equation: $y = \mathrm{Im} * a_1 (x - a_0)$

where,
PAR is y [µmol photons/m^2/s]
Digital signal is x [count]

Table 1 - Connector Pin Descriptions

Pin	Signal	Description
1	GND	Power Supply Return
2	RS-232 RX	RS-232 Receive
3	RS-232 TX	RS-232 Transmit
4	+V	Input Voltage
5	NC	Not Connected
6	NC	Not Connected

This is a non-standard pinout.



Male Face View

Satlantic
Richmond Terminal, Pier 9
3481 North Marginal Road
Halifax, Nova Scotia
B3K 5X8
Canada

Tel: 902-492-4780
Fax: 902-492-4781