

In Water Digital Linear PAR Sensor Calibration

Date: May 12, 2014

Customer: University of Washington

Project Number: 2014-7036 Model: PAR LIN 600m

Serial Number: 0507

Calibration Coefficients

Im = 1.3589

 $a_0 = 2157017270.0$ [count]

 $a_1 = 2.51157245562e-006 [\mu mol photons/m²/s/count]$

Equation: $y = \text{Im}^* a_1(x - a_0)$

where,

PAR is y [µmol photons/m²/s]

Digital signal is x [count]

Table 1 - Connector Pin Descriptions

Pin	Signal	Description
1	Power	Power
2	GND	Power Supply Return
3	NC	Not Connected
4	NC	Not Connected
5	TX	RS232
6	RX	RS232
7	NC	Not Connected
8	NC	Not Connected

Satlantic

Richmond Terminal, Pier 9 3481 North Marginal Road Halifax, Nova Scotia B3K 5X8 Canada

Tel: 902-492-4780 Fax: 902-492-4781

