

Certificate no: Oxygen Optode 4831_276_151488
FoilID: 1206EM

Product: **Oxygen Optode 4831**
Serial no: 276
Calibration date: 10.09.2018

2-POINT RECALIBRATION, MULTIPOINT CALIBRATED OXYGEN OPTODE

This is to certify that this product has been calibrated and verified using the following reference equipment:

Reference Equipment	Serial
Calibration Bath Model Toho TM-005-P-A	12-K
Fluke 5615 PRT	Serial No. 802054 and Serial No. 849155
Fluke CHUB E-4	Serial No. A7C677
Reference sensor 4330	339

Specification

	O2 Concentration	Air Saturation	Temperature
Range	0-500 µM	0 - 120%	-5 to +40°C
Accuracy	< ±3µM or ±2% ¹⁾	±2% ²⁾	±0.03°C ³⁾
Resolution	< 1 µM	< 0.4%	0.01°C
Settling Time (63%)	< 25 seconds	< 25 seconds	< 2 seconds

1) Requires salinity compensation for salinity variations > 1mS/cm, and pressure compensation for pressure > 100meter

2) Within calibrated range 0 - 120% / 0 - 30°C

3) Within calibrated range 0 - 36°C for standard version and 0 - 30°C for multipoint version

Calibration points and readings:

	Air Saturated Water	Zero Solution (Na2SO3)
Phase reading (°)	33.02	60.63
Temperature reading (°C)	9.98	22.17
Air Pressure (hPa)	1010.80	

Giving these coefficients

Index	0 (Offset)	1 (Slope)
ConcCoef	-2.436518E-01	1.067518E+00

With following settings

Index	0	1	2	3	4	5	6
SVUFoilCoef	2.79235E-03	1.16350E-04	2.36253E-06	2.29787E+02	-3.75842E-01	-5.60660E+01	4.57187E+00
PhaseCoef	0.000000E+00	1.000000E+00	0.000000E+00	0.000000E+00			
TempCoef	2.688810E+01	-3.114703E-02	3.170635E-06	-4.595761E-09	0.000000E+00	0.000000E+00	
Salinity (mS/cm)	0.00						

Date:10.09.2018



Sr. Repair Technician