



## SAMI<sup>2</sup> pH Validation Certificate

November 4, 2019

Serial Number: P0115 RMA Number: 082727 Technician: BW

Reagent: pmcp0574

Table 1: SAMI<sup>2</sup> pH E-values.

| Ea <sub>434</sub> | Eb <sub>434</sub> | Ea <sub>578</sub> | Eb <sub>578</sub> |
|-------------------|-------------------|-------------------|-------------------|
| 18103.0           | 2296.0            | 97.8              | 40427.0           |

## Validation Notes

Tris Batch: 101519  $\Delta T = 0.05 \,^{\circ}\text{C}$ 

Tris Bottle: 3 N=9

Validation Temp: 25.15 °C  $\Delta pH = -0.0023 \pm 0.0009$ 

## **Build Notes**

Pressure Test: 200 psi - Passed

 ${\bf Modifications:}$ 

Refurb with PV cable swap

5% reagent remain when received