



## SAMI<sup>2</sup> pH Validation Certificate

September 9, 2020

Serial Number: P0153 RMA Number: 082127 Technician: BW

Reagent: pmcp0714

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
17372.0	2284.0	94.1	38676.0

## Validation Notes

Tris Batch: 061720  $\Delta T = 0.03 \,^{\circ}\text{C}$ 

Tris Bottle: 9 N = 9

Validation Temp: 25.04 °C  $\Delta pH = -0.0001 \pm 0.0009$ 

## **Build Notes**

Pressure Test: 200 psi - Passed

 ${\bf Modifications:}$ 

Standard Refurbishment

Optical filters and fibers replaced.