



April 24, 2019

## **SAMI-pH Accuracy Certificate**

**Sunburst SAMI Serial Number: P0111**

**OOI Identification Number:**

**RSN P/N: 4830-58337**

**RSN AT: ATAPL58337-00003**

**Date: 11/12/13**

### **pH Evaluates:**

Ea<sub>434</sub> = 17533

Ea<sub>578</sub> = 101

Eb<sub>434</sub> = 2229

Eb<sub>578</sub> = 38502

### **Accuracy Test Temperature**

T = 25 C

Toffset = 0.05 C

**Certified Tris Batch #:** Tris 102913\_bag1

**Tris Bottle/Bag #:** Bag 1

**Tris Correction:** -0.0029

**mCP lot:** MKBC2604V

---

**Date: 10/22/13**

### **Pressure Test:**

**200psi – passed**

### **Modifications:**

- P/V lid with bubble trap
- Diaphragm retainer with ridge