



April 15, 2014

SAMI-pH Accuracy Certificate

Sunburst SAMI Serial Number: P0132

OOI Identification Number:

CGSN P/N: 1336-00013-000004

PHSEN Series D – PH3304

Date: 4/15/14

pH Evaluates:

$Ea_{434} = 17533$

$Ea_{578} = 101$

$Eb_{434} = 2229$

$Eb_{578} = 38502$

Accuracy Test Temperature

$T = 25\text{ C}$

$T_{offset} = 0\text{ C}$

Certified Tris Batch #: 032714

Tris Bottle/Bag #: bottle

Tris Correction: -0.005

mCP lot: MKBC2604V

Date: 4/14/14

Pressure Test:

200psi – passed

Modifications:

Pump/valve bubble trap

Modified diaphragm retainer ring