



## $\mathrm{SAMI}^2$ pH Validation Certificate

October 18, 2024

Serial Number: P0113 RMA Number: 784 Technician: BW Reagent: 1388

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
19168.0	2546.0	111.0	41962.0

## Validation Notes

Tris Batch: 071524  $$\Delta\,{\rm T}=0.15\,^{\circ}{\rm C}$$ 

Tris Bottle: 7 N=7

Validation Temp: 24.92 °C  $\Delta pH = 0.0019 \pm 0.0009$ 

## **Build Notes**

Pressure Test: 200 psi – Passed

 ${\bf Modifications:}$ 

standard refurb 2x sem filters IR testing