



SAMI² pH Validation Certificate

November 11, 2019

Serial Number: P0163 RMA Number: 082720

Technician:

Reagent: pmcp0582

Table 1: $SAMI^2$ pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
17372.0	2284.0	94.1	38676.5

Validation Notes

Tris Batch: 101519 $\Delta T = 0.07 \,^{\circ}\text{C}$

Tris Bottle: 4 N=7

Validation Temp: 25.20 °C $\Delta pH = -0.0026 \pm 0.0011$

Build Notes

Pressure Test: 200 psi – Passed

Modifications:

refurbishment with OOI 3-pin upgrade; pv cable, bulkhead, pressure plate and labor.

Replacement of pump and valve

75% reagent remained on rcpt