



In Water Digital Linear PAR Sensor Calibration

Date: December 18, 2013
 Customer: Oregon State University
 Project Number: 2013-7020
 Model: PAR LIN 600m
 Serial Number: 434

Calibration Coefficients


$I_m = 1.3589$
 $a_0 = 2156561751.4$ [count]
 $a_1 = 2.5063877597725e-006$ [$\mu\text{mol photons/m}^2/\text{s/count}$]

Equation: $y = I_m * a_1 (x - a_0)$

where,
 PAR is y [$\mu\text{mol photons/m}^2/\text{s}$]
 Digital signal is x [count]

Table 1 - Connector Pin Descriptions

Pin	Signal	Description
1	Power	Power
2	GND	Power Supply Return
3	NC	Not Connected
4	NC	Not Connected
5	TX	RS232
6	RX	RS232



Satlantic
 Richmond Terminal, Pier 9
 3481 North Marginal Road
 Halifax, Nova Scotia
 B3K 5X8
 Canada

Tel: 902-492-4780
 Fax: 902-492-4781