



April 24, 2019

SAMI-pH Accuracy Certificate

Sunburst SAMI Serial Number: P0112

OOI Identification Number:

RSN P/N: 4830-58337

RSN AT: ATAPL58337-00004

Date: 11/7/13

pH Evaluates:

$Ea_{434} = 17533$

$Ea_{578} = 101$

$Eb_{434} = 2229$

$Eb_{578} = 38502$

Accuracy Test Temperature

$T = 25\text{ C}$

$\text{Toffset} = 0.07\text{ C}$

Certified Tris Batch #: Tris 102913_bag1

Tris Bottle/Bag #: Bag 1

Tris Correction: -0.0028

mCP lot: MKBC2604V

Date: 10/22/13

Pressure Test:

200psi – passed

Modifications:

- P/V lid with bubble trap
- Diaphragm retainer with ridge