

Certificate no: Oxygen Optode 4831\_275\_135865  
FoillID: 1206EM

Product: **Oxygen Optode 4831**  
Serial no: 275  
Calibration date: 23.10.2017

## 2-POINT RECALIBRATION, MULTIPOINT CALIBRATED OXYGEN OPTODE

This is to certify that this product has been calibrated and verified using the following reference equipment:

Reference Equipment	Serial
Calibration Bath Model Toho TM-005-P-A	12-K
Fluke 5615 PRT	Serial No. 802054 and Serial No. 849155
Fluke CHUB E-4	Serial No. A7C677
Reference sensor 4330	338

### Specification

	O2 Concentration	Air Saturation	Temperature
Range	0-500 $\mu$ M	0 - 120%	-5 to +40°C
Accuracy	< $\pm 3\mu$ M or $\pm 2\%$ <sup>1)</sup>	$\pm 2\%$ <sup>2)</sup>	$\pm 0.03^\circ$ C <sup>3)</sup>
Resolution	< 1 $\mu$ M	< 0.4%	0.01°C
Settling Time (63%)	< 25 seconds	< 25 seconds	< 2 seconds

1) Requires salinity compensation for salinity variations > 1mS/cm, and pressure compensation for pressure > 100meter

2) Within calibrated range 0 - 120% / 0 - 30°C

3) Within calibrated range 0 - 36°C for standard version and 0 - 30°C for multipoint version

### Calibration points and readings:

	Air Saturated Water	Zero Solution (Na2SO3)
Phase reading (°)	34.01	60.40
Temperature reading (°C)	10.05	21.65
Air Pressure (hPa)	988.00	

### Giving these coefficients

Index	0 (Offset)	1 (Slope)
ConcCoef	-4.055371E+00	1.093039E+00

### With following settings

Index	0	1	2	3	4	5	6
SVUFoilCoef	2.81240E-03	1.18234E-04	2.47037E-06	2.28712E+02	-3.48520E-01	-5.90578E+01	4.55274E+00
PhaseCoef	0.000000E+00	1.000000E+00	0.000000E+00	0.000000E+00			
TempCoef	2.561202E+01	-3.076118E-02	2.920578E-06	-4.349522E-09	0.000000E+00	0.000000E+00	
Salinity (mS/cm)	0.00						

Date:23.10.2017



Sr. Repair Technician