



## SAMI<sup>2</sup> pH Validation Certificate

February 3, 2017

Serial Number:

P0059

OOI Number:

Technician: B. Wasser

Reagent: pmcp0127

Table 1: SAMI<sup>2</sup> pH E-values.

$Ea_{434}$	$\mathrm{Eb_{434}}$	$\mathrm{Ea}_{578}$	$\mathrm{Eb}_{578}$
17372.0	2284.1	94.1	38676.5

## Validation Notes

Tris Batch: 111616

 $\Delta T = 0.10 \, ^{\circ}C$ 

Tris Bottle: 6

N = 9

Validation Temp: 24.88 °C

 $\Delta \, \mathrm{pH} = 0.0003 \pm 0.0011$ 

## **Build Notes**

Pressure Test: 200 psi-Passed

Modifications:

Swagelok Plug

Peek Mixing tube