



SAMI² pH Validation Certificate

October 7, 2020

Serial Number: P0132 RMA Number: 083125

> Technician: B. Montgomery Reagent: PmCP_0725

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
17372.0	2284.0	94.1	38676.0

Validation Notes

Tris Batch: 091620 $\Delta T = 0.10 \,^{\circ}\text{C}$

Tris Bottle: 3 N=7

Validation Temp: 25.42 °C $\Delta pH = -0.0020 \pm 0.0023$

Build Notes

Pressure Test: 200 psi - Passed

 ${\bf Modifications:}$

Standard refurbishment

3-pin bulkhead replaced with used part

6-pin bulkhead replaced

Diaphragm replaced