

Sunburst Sensors, LLC 1226 W Broadway Ave Missoula, MT 59802 +1 406 532 3247

$SAMI^2$ CO_2 Calibration Certificate

January 11, 2017

Serial Number:

C47

Technician:

B. Wasser

Reagent:

btbed0022

Table 1: $SAMI^2 CO_2$ calibration parameters.

·A	В	\mathbf{C}	K_{434}	K_{620}	T_{avg}	Range (ppm)
0.0660	0.5863	-1.5202	1.0441	0.9722	4.7005	203-1198

Table 2: Blank intensities.

Reference	Signal	
41	44	
2672	2790	
2786	2713	
	41 2672	

Build Notes

Pressure Test: 200 psi

Modifications:

Diaphragm upgrade Col Cal, Ext Range

