



SAMI² pH Validation Certificate

April 2, 2024

Serial Number: P0116 RMA Number: 281 Technician: BW

Reagent: pmcp1278

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb_{578}
17372.0	2284.0	94.1	38676.5

Validation Notes

Tris Batch: 013214 $\Delta T = 0.17$ °C

Tris Bottle: 9 N=9

Validation Temp: 24.98 °C $\Delta pH = -0.0015 \pm 0.0005$

Build Notes

Pressure Test: 200 psi - Passed

 ${\bf Modifications:}$

standard refurbishment

IR testing

pump replaced

2x semrock filter upgrade

bulkhead replaced with cust prov. part 1/2 labor