

${\rm SAMI^2~CO_2~Calibration~Certificate}$

November 22, 2021

Serial Number: c79

RMA Number:

OOI Number: 091311 Technician: BW Reagent: btb0408

Table 1: $SAMI^2 CO_2$ calibration parameters.

A	В	$^{\mathrm{C}}$	K_{434}	K_{620}	T_{avg}	Range (ppm)
-0.0131	0.9882	-2.2834	0.8807	1.3519	12.7140	280-1197

Table 2: Blank intensities.

Source	Reference	Signal
Dark	88	105
$434~\mathrm{nm}$	3180	2830
$620~\mathrm{nm}$	2255	3032

Build Notes

Pressure Test: 200 psi

Modifications:

Standard Refurbishment 434 filter replaced extended range calibration

IR testing

