



SAMI² pH Validation Certificate

September 20, 2020

Serial Number: P0161 RMA Number: 082127

> Technician: B. Montgomery Reagent: PmCP_0715

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
17372.0	2284.0	94.1	38676.0

Validation Notes

Tris Batch: 091620 $\Delta T = 0.09$ °C

Tris Bottle: 1 N=7

Validation Temp: 25.27 °C $\Delta pH = 0.0016 \pm 0.0007$

Build Notes

Pressure Test: 200 psi – Passed

 ${\bf Modifications:}$

Standard refurbishment

Fibers replaced