



SAMI^2 pH Validation Certificate

July 21, 2023

Serial Number: P131 RMA Number: 082307 Technician: BP

Reagent: PmCP119

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
17372.0	2284.0	94.1	38676.5

Validation Notes

Tris Batch: 071823 $\Delta T = 0.16\,^{\circ}\text{C}$

Tris Bottle: N = 9

Validation Temp: 25.04 °C $\Delta pH = 0.0008 \pm 0.0026$

Build Notes

Pressure Test: 200 psi – Passed

 ${\bf Modifications:}$

Stadard Refurb, new cell, new filters