

$SAMI^2 CO_2$ Calibration Certificate

May 16, 2023

Serial Number: C0076 RMA Number: 082343 Technician: BW

Reagent: btb_0501

Table 1: SAMI² CO₂ calibration parameters.

A	В	\mathbf{C}	K_{434}	K_{620}	T_{avg}	Range (ppm)
0.0449	0.6318	-1.5510	0.8147	1.2438	10.6596	203-1187

Table 2: Blank intensities.

Source	Reference	Signal
Dark	64	50
$434~\mathrm{nm}$	3095	2518
$620~\mathrm{nm}$	1815	2228

Build Notes

Pressure Test: 200 psi

Modifications:

0

Standard Refurbishment

IR Testing + replacements

new pump- low flow

swapped board, pad damaged-no comms

10 pct Reagent

Ext Cal

