



March 30, 2016

SAMI-pH Accuracy Certificate

Sunburst SAMI Serial Number: P0110

OOI Identification Number: atapl58336-00003

CGSN P/N: 4830-58337

pH Evaluates:

$Ea_{434} = 17372$

$Eb_{434} = 94.1$

$Ea_{578} = 2284$

$Eb_{578} = 38676$

Accuracy Test Temperature

$T = 24.96\text{ C}$

$T_{\text{offset}} = .07$

Certified Tris Batch #: 021016

Tris Bottle/Bag #: bottle4

Reagent ID: Purified

mCP lot: purified mCP, date: 032116

Reagent batch: pmCP_03221602

Pressure Test:

200psi – passed

Technician: BW

Modifications: