



## $SAMI^2$ pH Validation Certificate

November 12, 2021

Serial Number: P0113 RMA Number: 091340

> Technician: B. Montgomery Reagent: PmCP\_0864

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
17372.0	2284.0	94.1	38676.0

## Validation Notes

Tris Batch: 092921  $\Delta T = 0.00 \,^{\circ}\text{C}$ 

Tris Bottle: 5 N=9

Validation Temp: 24.78 °C  $\Delta pH = -0.0006 \pm 0.0010$ 

## **Build Notes**

Pressure Test: 200 psi - Passed

Modifications:

Standard refurbishment

Insulation resistance testing

Board replaced (used)

Battery chassis replaced

Backup batteries and harness replaced

Fibers replaced

434, 578 and ND filters replaced