



Sunburst Sensors, LLC  
1226 W Broadway Ave  
Missoula, MT 59802  
+1 406 532 3246

## SAMI<sup>2</sup> pH Validation Certificate

November 4, 2019

Serial Number: P0164  
RMA Number: 082727  
Technician: BW  
Reagent: pmcp0575

Table 1: SAMI<sup>2</sup> pH E-values.

Ea <sub>434</sub>	Eb <sub>434</sub>	Ea <sub>578</sub>	Eb <sub>578</sub>
18103.0	2296.0	97.8	40427.0

## Validation Notes

Tris Batch: 101519

$\Delta T = 0.07^{\circ}\text{C}$

Tris Bottle: 3

N = 9

Validation Temp: 25.10 °C

$\Delta \text{pH} = -0.0054 \pm 0.0014$

## Build Notes

Pressure Test: 200 psi – Passed

Modifications:

refurb with PV cable replacement