

${\rm SAMI^2~CO_2~Calibration~Certificate}$

December 11, 2023

Serial Number: c79

RMA Number:

OOI Number: 276 Technician: BW

Reagent: btb_0534

Table 1: $SAMI^2 CO_2$ calibration parameters.

A	В	С	K_{434}	K_{620}	T_{avg}	Range (ppm)
0.0495	0.6194	-1.5153	0.9778	2.0147	7.3871	202-1188

Table 2: Blank intensities.

Source	Reference	Signal
Dark	84	102
$434~\mathrm{nm}$	2939	2892
$620~\mathrm{nm}$	1585	3130

Build Notes

Pressure Test: 200 psi

Modifications:

standard refurb

IR testing

ext cal

