



October 30, 2020

OOI SAMI-CO₂ Calibration Certificate

Sunburst SAMI Serial Number: C0078

OOI Identification #:

RSN P/N: 4830-58336

PCO₂W Series A – PC2301

Date: 12/18/13

Calibration Range (ppm): 216-1158

CO₂ Calibration Coefficients:

a = 0.0080

b = 0.7613

c = -1.6393

K₄₃₄ = 0.7735

K₆₂₀ = 1.0485

CO₂ Calibration Temperature

T = 14.94

Pressure Test:

200 psi

Modifications:

Pump/valve lid with bubble trap

Diaphragm retainer with ridge