



SAMI² pH Validation Certificate

October 25, 2024

Serial Number: P0111 RMA Number: 784 Technician: BW Reagent: 1397

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
19168.0	2546.0	111.0	41962.0

Validation Notes

Tris Batch: 071524 $\Delta T = 0.10$ °C

Tris Bottle: 9 N=9

Validation Temp: 25.00 °C $\Delta pH = -0.0038 \pm 0.0007$

Build Notes

Pressure Test: 200 psi – Passed

 ${\bf Modifications:}$

Standard Refurb

IR testing

2x semrock filter upgrade

15% reagent remaining at rcpt