



## In Water Digital Linear PAR Sensor Calibration

Date: January 11, 2016  
 Customer: University of Washington  
 Project Number: 2015-3273  
 Model: PAR LIN 600m  
 Serial Number: 463

---

### Calibration Coefficients

$I_m = 1.3589$   
 $a_0 = 2156563217$  [count]  
 $a_1 = 2.425271e-006$  [ $\mu\text{mol photons/m}^2/\text{s/count}$ ]

---


Equation:  $y = I_m * a_1 (x - a_0)$

where,  
 PAR is  $y$  [ $\mu\text{mol photons/m}^2/\text{s}$ ]  
 Digital signal is  $x$  [count]

---

**Table 1 - Connector Pin Descriptions**

Pin	Signal	Description
1	Power	Power
2	GND	Power Supply Return
3	NC	Not Connected
4	NC	Not Connected
5	TX	RS232
6	RX	RS232



**Satlantic**  
 Richmond Terminal, Pier 9  
 3481 North Marginal Road  
 Halifax, Nova Scotia  
 B3K 5X8  
 Canada

Tel: 902-492-4780  
 Fax: 902-492-4781