



SAMI² pH Validation Certificate

November 2, 2023

Serial Number: P0110 RMA Number: 280 Technician: BW

Reagent: pmcp1189

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
17372.0	2284.0	94.1	38676.0

Validation Notes

Tris Batch: 121223 $\Delta T = 0.16\,^{\circ}\text{C}$

Tris Bottle: 3 N=9

Validation Temp: 25.00 °C Δ pH = 0.0011 ± 0.0025

Build Notes

Pressure Test: 200 psi - Passed

 ${\bf Modifications:}$

standard refurbishment

3 pim BH, 3-pin cable, pressure cap, thermistor

434 + 578 Sem u/g