



Sunburst Sensors, LLC
1226 W Broadway Ave
Missoula, MT 59802
+1 406 532 3247

SAMI² pH Validation Certificate

February 3, 2017

Serial Number: P0111
Technician: B. Wasser
Reagent: pmcp0125

Table 1: SAMI² pH E-values.

Ea ₄₃₄	Eb ₄₃₄	Ea ₅₇₈	Eb ₅₇₈
17372.0	2284.1	94.1	38676.5

Validation Notes

Tris Batch: 111616

$\Delta T = 0.07^{\circ}\text{C}$

Tris Bottle: 6

$N = 9$

Validation Temp: 24.89 °C

$\Delta \text{pH} = -0.0055 \pm 0.0045$

Build Notes

Pressure Test: 200 psi – Passed

Modifications:

Swagelok Plug

Peek Mixer tube