



Certificate of Calibration


Ocean Sonics, Ltd.

Calibration Certificate Number: **C9015**

Test Result: 10 kHz to 100 kHz: -173.9 ± 2.5
10 kHz to 200 kHz: -173.7 ± 3.1

Model Number SB35-ETH
Serial Number 1362
Manufacture Date 7-May-2014
Measurement Date 6-Nov-2024
Certificate Date 6-Nov-2024
Sensitivity @ 26 Hz $-170.5 \text{ dB re V / uPa}$
Case Type Titanium
Element Manufacturer GeoSpectrum
Element Model M24HF
Element Serial F000365
Preamp Model 04-300435-09
Calibrated By D.Lee
Work Order Number W45-135
Test Type RX Sensitivity
Test Procedure Complex RMS
Test Location Tank #3, 1 m
Water Temperature 16 °C

Projector Manufacturer Ocean Sonics
Projector Model TH2-SER-4F
Projector Serial 2225
Measurement Distance 1 m
Output Level 130.6 dB re uPa @ 1 m
Tone Burst 100.0 us / 300 ms
Reference Manufacturer Ocean Sonics
Reference Model RB9-ETH
Reference Serial 6465
Primary Calibration 2023-07-04
Preamp Manufacturer Ocean Sonics
Preamp Model 04-300451-01
Preamp Serial Z88
Preamp Gain 36 dB
ADC Manufacturer Ocean Sonics
ADC Model Number 04-300427-01
ADC Serial Number Z87



Frequency kHz	Sensitivity [dBV re uPa]	
	0 deg	90 deg
10.0	-171.9	-172.0
20.1	-172.3	-173.0
30.1	-173.3	-173.3
40.2	-172.5	-172.3
50.2	-172.8	-173.3
60.2	-171.8	-171.4
70.3	-171.5	-171.5
80.3	-174.6	-173.5
90.4	-176.5	-175.4
100.4	-175.9	-174.8
110.4	-176.8	-173.9
120.5	-174.8	-173.2
130.5	-172.2	-171.8
140.5	-171.3	-171.1
150.6	-170.7	-170.7
160.6	-171.2	-171.7
170.7	-171.0	-170.5
180.7	-171.7	-171.0
190.7	-173.4	-172.7
200.0	-176.0	-175.0

#	Quantity	Part	Issue	Type	Resolution
4	1 ea	SB35-ETH - icListen HF - Ethernet, 3500m Depth Rating, 10 Hz - 200 kHz. Includes case, test cable and power adapter, with 256 GB memory. Country/Origin: Canada SN(s): 1250;		Repair	

US Shipments

If you need to inform the ECCN code please use the following:

- * The ECCN code for deep hydrophones (more than 1000m) is 6A001.a.2.a.6
- * The ECCN code for our 200m and 900m (less than 1000m) hydrophones is EAR99

General Shipping Instructions for the Return of **RMA#757** to Ocean Sonics:

1. **Please, when shipping the goods back to Canada use the HS code 901590 for the accessories and HS code 901580 for the hydrophones.. State clearly on the airway bill, customs declaration form and commercial invoice, "NOT FOR SALE! Canadian Goods Returning for Repair. Value is for Customs Purposes Only." and the RMA #757.

Important: If you are shipping more than 2 hydrophones, declare when filling your air way bill: Ion lithium battery contained in equipment (UN3481, PI 967, section II).

Some courier companies have this drop-down option. If there is you must choose it. If there is not, you have to write this statement in the comment section of after the description of the goods.

Outside the box place two UN3481 labels, one in each opposite side, and with a marker, write near the label: 'PI 967 - II'

2. Print and add the RMA document to the box with the returning unit.

3. When filling the airway and the Commercial Invoice for the return please add the following: NOTIFY:

PF Collins
251 E White Hills Rd #100, St. John's, NL A1A 5X7
Canada
Email: import@pfcollins.com , Phone: +1 709 726 7596

NOTE: If the courier company has the option, you can use the Broker Select Option (BSO). This way our broker will be notified upon arrival in Canada.

4. Guideline for valuation of returned equipment: The value for cross border shipment is used for transport insurance and customs value. It should represent the replacement cost of the equipment (to same standard equipment). We suggest 100% of the equipment value.

5. By sending the RMA I acknowledge that the item(s)listed herein are returned without having been advanced in value or improved in condition by any process of manufacture or other means. I acknowledge that all repairs are subject to the Ocean Sonics Terms and Conditions.