



Sunburst Sensors, LLC

1226 W Broadway Ave

Missoula, MT 59802

+1 406 532 3247

SAMI2 CO2 Calibration Certificate

Serial Number: C0064

Board Version: E

Technician: BW

RMA #: 1487

Reagent Batch: 670

Reagent Remaining: 10%

Calibration Date: 2026-03-30

Notes:

Standard refurb

IR test- pass

2x SEM filters

Ext Cal

Table 1: SAMI2 CO2 calibration parameters

A:	0.079744
B:	0.469013
C:	-1.40545
k434:	1.1955
k620:	1.69489
Tavg (°C):	9.94686
Range (μ atm):	200 – 2000

Table 2: Representative blank intensities

Dark Ref: 18
Dark Sig: 43
434 Ref: 3038
434 Sig: 3660
620 Ref: 2017
620 Sig: 3427

