



Sunburst Sensors, LLC

1226 W Broadway Ave

Missoula, MT 59802

+1 406 532 3247

SAMI2 CO2 Calibration Certificate

Serial Number: C0080

Board Version: E

Technician: BW

RMA #: 1487

Reagent Batch: 671

Reagent Remaining: 75%

Calibration Date: 2026-03-30

Notes:

Standard Refurbishment - flooded - Replaced board w/ Used E-Board
Ext Cal.

2x SEM filters

IR Test - Fail - Replaced Bulkheads w/ cust supplied part

New Pressure cap

Table 1: SAMI2 CO2 calibration parameters

A:	0.0484176
B:	0.623546
C:	-1.5847
k434:	1.28064
k620:	1.04541
Tavg (°C):	9.98796
Range (μ atm):	200 – 2200

Table 2: Representative blank intensities

Dark Ref: 26
Dark Sig: 79
434 Ref: 2352
434 Sig: 3075
620 Ref: 3086
620 Sig: 3289

