



# VistA Blood Establishment Computer Software (VBECS) – Echo Interface Configuration and Setup Guide, Version 4.0

September 2016

Department of Veterans Affairs  
Enterprise Project Management Office

## Revision History

Date	Revision	Description	Author
06-03-16	1.0	Initial version.	BBM team
		Defect 337273 <ul style="list-style-type: none"> <li>Figure 4 Corrected driver name</li> </ul> Section 4.1 <ul style="list-style-type: none"> <li>Added information bar about using local procedures to download configuration files</li> <li>Removed Step 1 that had the user getting the configuration files directly from the Instrument Manager server</li> <li>Added step 5 to copy files to the Instrument Manager Server</li> <li>Updated step 6 to say on the Instrument Manager Server</li> <li>Added an information bar about copying and pasting into a PowerShell window</li> <li>Updated Figure 8 with new checksums</li> </ul> Section 4.3 <ul style="list-style-type: none"> <li>Changed Verify step to use a report instead of screen captures</li> <li>Figure 13 Updated to show Results Test Code Mapping Report Window</li> </ul> Section 4.4 <ul style="list-style-type: none"> <li>Changed Verify step to use a report instead of screen captures</li> <li>Figure 14: Updated to show Rules Report Window</li> </ul> Section 4.7 <ul style="list-style-type: none"> <li>Changed Verify step to use a report instead of screen captures</li> <li>Figure 25: Updated to show Results Test Code Mapping Report Window</li> </ul> Section 5 <ul style="list-style-type: none"> <li>Changed steps 9 and 10 to include save steps</li> <li>Removed step 12</li> <li>Appendix A: Reordered alphabetically by Instrument Test Code</li> <li>Appendix B: Reordered alphabetically by Instrument Test Code</li> <li>Appendix C: Changed to use the report instead of screen captures</li> <li>Appendix G: Updated validation scenarios</li> </ul> Updated Footer to include use of the Electronic Document Control System.	
06-28-16	2.0	Moved Footer notes to Header, Updated Document to Version to 2.0	BBM team
		Defect 355473: <ul style="list-style-type: none"> <li>Added warning about discrepancy in driver versions in sections 4.2 and 4.6</li> <li>Added information about proper configuration of test Specimen UIDs to Test Group Two warning flag section</li> <li>Removed information about Event Log from Test Group One and replaced it with checking Audit Trail Report</li> </ul>	
08-19-16	3.0		BBM team
		Defect 372140: <ul style="list-style-type: none"> <li>Set Up Automated Instrument, Testing Template Setup section: Added informational message that refers to AABB standards for crossmatch</li> <li>Updated names of configuration files and their checksums in section 4</li> <li>Updated Figures 18 and 20 to show the new rule</li> <li>Added new rule to Appendix C</li> </ul>	
09-08-16	4.0		BBM team

## Table of Contents


REVISION HISTORY.....	2
TABLE OF CONTENTS.....	3
INTRODUCTION .....	5
RELATED MANUALS AND REFERENCE MATERIALS .....	6
SET UP AUTOMATED INSTRUMENT .....	8
1    Network Connectivity Setup .....	8
2    Testing Template Setup.....	8
3    User ID Setup .....	8
SET UP INSTRUMENT MANAGER .....	9
1    Instrument Manger version.....	9
2    Instrument Manager to Automated Instrument Connectivity .....	9
3    Installing Instrument Driver .....	10
4    Set up Instrument Side Configuration .....	10
4.1    Download instrument configuration files .....	10
4.2    Import Instrument side configuration .....	12
4.3    Verify test code mapping for instrument side configuration .....	15
4.4    Verify rules for instrument side configuration.....	17
4.5    Configure rules for instrument side configuration.....	18
4.6    Import VBECS (HL7) side configuration.....	21
4.7    Verify test code mapping for VBECS side configuration.....	24
4.8    Verify rules for VBECS (HL7) side configuration.....	25
5    Set Up HL7 Connection to VBECS TEST .....	28
6    Set Up Instrument Connection .....	30
7    Test New Connections.....	32
8    Validate Instrument connectivity to VBECS TEST .....	33
9    Set up HL7 Connection to VBECS PROD.....	33
GLOSSARY .....	38
APPENDICES .....	39
APPENDIX A: INSTRUMENT SIDE MAPPING .....	39
APPENDIX B: HL7 (VBECS) SIDE MAPPING.....	40
APPENDIX C: INSTRUMENT SIDE RULES.....	41
APPENDIX D: VBECS (HL7) SIDE RULES .....	43
APPENDIX E: ECHO TESTING TEMPLATES .....	44
APPENDIX F: CA SDM SUPPORT TICKET TEMPLATE.....	46
APPENDIX G: VALIDATION PLANNING AND EXAMPLE TEST SCENARIOS .....	47
TEST GROUP ONE: AI INTERFACE DISABLED .....	48
TEST GROUP TWO: VERIFY AI INDIVIDUAL TEST(S).....	49
TEST GROUP 2 SCENARIO 1: VERIFY AI TAS TEST .....	50
TEST GROUP 2 SCENARIO 2: VERIFY AI SEROLOGIC CROSSMATCH TEST .....	51
TEST GROUP 2 SCENARIO 3: VERIFY AI PATIENT DIAGNOSTIC TESTS.....	52
TEST GROUP 2 SCENARIO 4: VERIFY AI BLOOD UNIT TESTS .....	53


This page intentionally left blank.


## Introduction

VBECS is a Blood Bank application that facilitates ongoing compliance with Food and Drug Administration (FDA) standards for medical devices and enhances the VA's ability to produce high-quality blood products and services to veterans. The system follows blood bank standards, standards of national accrediting agencies, FDA regulations and VA policies.

VBECS 2.2.0 introduced a new interface for blood bank testing performed by blood bank instrumentation to VBECS. The implementation of the interface and its associated validation are described in this guide.

 *Unauthorized access or misuse of this system and/or its data is a federal crime. Use of all data must be in accordance with VA security and privacy policies.*


 *The U.S. FDA classifies this software as a medical device. Unauthorized modifications will render this device an adulterated medical device under Section 501 of the Medical Device Amendments to the Federal Food, Drug, and Cosmetic Act. Acquiring and implementing this software through the Freedom of Information Act requires the implementer to assume total responsibility for the software and become a registered manufacturer of a medical device, subject to FDA regulations. Adding to or updating VBECS software without permission is prohibited.*

 *Instructions in this Setup Guide must be followed for the interface to deliver information to VBECS. Local validation is required to confirm proper operation before use. Validation and verification is required to ensure connectivity to VBECS.*

This guide is provided to assist you with the multi-faceted required setup of your local blood bank testing instrument(s), Data Innovations Instrument Manager (DI IM) and VBECS to electronically transmit instrument test results to VBECS for use in the transfusion service.

There are specific setup requirements to test and transmit those testing results to VBECS for review using DI IM (**Figure 1**).

**Figure 1: Hardware and Interface Configuration**




Your local testing instrument(s) communicates with DI IM via an instrument specific driver provided by DI that must be downloaded from DI and installed, locally.

DI IM communicates directly with VBECS via a generic HL7 interface driver that must be downloaded from DI and installed, locally. This driver is then customized for VBECS by downloading and installing the driver configuration file.

VBECS has an interface that must be configured in VBECS Administrator to receive messages from DI IM (**Figure 2**).

**Figure 2: Setup Path**



## Related Manuals and Reference Materials

- *VistA Blood Establishment Computer Software (VBECS) 2.2.0 Technical Manual-Security Guide*
- *Data Innovations Instrument Manager Manual*
- *Blood Bank Analyzer User's Guide (Instrument Manual)*

This page intentionally left blank.

# Set Up Automated Instrument

## 1 Network Connectivity Setup

In order to ensure a proper functioning of an interface between an Automated Instrument and Instrument Manager the Instrument needs to be connected to the VA network. The static IP Address and Port number has to be assigned to the Instrument (further referred in this document as **<Instrument IP>** and **<Instrument Port>**). Please refer to the Instrument Manual or contact your vendor for the instructions about how to perform this setup. Check with Local IT staff to establish the connection to Instrument Manager.

## 2 Testing Template Setup



*Test results sent using templates that are not supported will not be accepted in VBECS.*

Please refer to **Appendix E: Echo Testing Templates** for the list of supported instrument templates.



*AABB Standards require a serological XM to detect ABO incompatibility (5.16.1) and a local policy must be in place if the site is not performing an IS AHG as part of their serologic crossmatch test, manual or using an instrument.*

## 3 User ID Setup



*Failure to set up proper user IDs on an instrument will prevent instrument results from being accepted in VBECS.*

In order for VBECS to properly recognize the person who performed testing on an Instrument, all users accessing the instrument must have their user IDs set up to match their network user IDs (e.g., VHAISHBURNISK).

Please refer to the Instrument Manual or contact your vendor for the instructions about how to perform setup of user IDs.



## Set up Instrument Manager

### 1 Instrument Manger version

Please verify that you are using Instrument Manager Version 8.13 or greater (Figure 3). In Instrument Manager, navigate to **Help -> About Instrument Manager**.

**Figure 3: Instrument Manager Version Screen**



*If your version of Instrument Manger is older than 8.13 please **STOP** executing this guide and update the software first. **Do not proceed until the issue is resolved.***



*If your Instrument Manager is greater than 8.13 you may proceed. The user must execute the instructions and validate functionality on newer version. Discrepancies in the instructions must be reported as a CA-SDM ticket. See **Appendix F**.*

### 2 Instrument Manager to Automated Instrument Connectivity

Please contact your local network administrative staff and ensure that your local network allows two-way TCP/IP connectivity between <**Instrument Manager IP**> address and <**Instrument IP**> address on <**Instrument Port**>.

### 3 Installing Instrument Driver

Two drivers are required for the correct operation of the Automated Instrument interface to VBECS (Figure 4):

**Figure 4: Required Drivers**

Driver	Name	Version	Description
Immure Galileo Echo/Arkonet	ibggecoi	8.00.0007	Instrument to DI IM driver
Data Innovations LLC, Configurable HL7	diihl7ml	8.00.0044	DI IM to VBECS (HL7) driver

*Please refer to the user's manual for Instrument Manager or contact Data Innovations for the instructions on how to install drivers for Instrument Manager.*

After successful installation of drivers please go to **Report -> Available Drivers** menu option in Instrument Manager and verify that the drivers listed in Figure 4 are present.

### 4 Set up Instrument Side Configuration



**Execute instructions in this section for each Echo instrument that will be connecting to VBECS.**



***Modifying rules or test code mappings as imported using this Instrument Manager configuration Setup Guide may lead to malfunction of the Automated Instrument to VBECS interface.***

Prerequisites for the Instrument Manager Configuration files download:

- You must be an administrator on *<Instrument Manager Server>*.
- Once the above prerequisite is met you may proceed.


#### 4.1 Download instrument configuration files



**Use local procedures for copying the instrument configuration files to the Instrument Manager server.**


1. Navigate to **<http://vaww.oed.portal.va.gov/projects/vbecs/Pages/Instrument-Manager-Configuration-Files-Release.aspx>** (Figure 5).

**Figure 5: Connecting to the VBECS Share**




2. To download a file from the SharePoint, right-click on it and select **Save target as** (Figure 6).

**Figure 6: Example of Save target as...**



3. In the next screen, specify the directory to save (Figure 7).

**Figure 7: Example of Save As**



4. Save both the **Echo – HL7 Interface Configuration Template for VBECS 09082016.gsb** and the **Echo – Instrument Interface Configuration Template for VBECS 09082016.gsb** files to the **C:\temp** directory.

This document is version controlled via approval signature. Do not assume that a printed, copied, or downloaded document is current. Consult the Electronic Document Control System to verify use of the current version.

5. Per local procedures copy both files to **C:\temp** on the Instrument Manager server
6. On the Instrument Manager server; Click **Start**, and in the “**Search programs and files**” box type **Run** and hit enter. Type **powershell** and click **OK** to launch PowerShell
7. Copy and paste or type the following commands to generate checksums for configuration files:

```
certutil -hashfile "C:\Temp\Echo - HL7 Interface Configuration Template for VBECS 09082016.gsb" MD5 <press Enter>
```

```
certutil -hashfile "C:\Temp\Echo - Instrument Interface Configuration Template for VBECS 09082016.gsb" MD5 <press Enter>
```



To copy, highlight the lines in grey and enter CTRL-C.

To insert the copied line into a PowerShell window, right click in the PowerShell window and select “Paste”.

8. Verify that checksums for both files match those shown in Figure 8.

**Figure 8: Instrument Manager Configuration File checksums**

```
Administrator: Windows PowerShell
PS C:\> certutil -hashfile "C:\Temp\Echo - HL7 Interface Configuration Template for VBECS 09082016.gsb" MD5
13 24 af 32 66 a2 db a0 47 f7 41 66 3c 1b 5f 88
PS C:\> certutil -hashfile "C:\Temp\Echo - Instrument Interface Configuration Template for VBECS 09082016.gsb" MD5
27 03 9b eb bd 4b 31 31 94 8f 37 90 6b 59 e4 72
certutil -hashfile command completed successfully.
PS C:\>
```




*If the checksums do not match, stop and file a national CA SDM ticket to coordinate assistance with installation using the template in Appendix F. Do not proceed until the issue is resolved.*

9. Close the **PowerShell** window.

## 4.2 Import Instrument side configuration


1. After downloading and verifying configuration files, open Instrument Manager and navigate to **Configuration -> Configuration Editor**.
2. Click on the **Import** button (Figure 9).

**Figure 9: Example of Configuration Editor**



3. Once the import screen opens, click the **Browse** button and select **C:\Temp** folder (Figure 10).
4. Click OK.

**Figure 10: Example of Configuration Import Screen**



5. Select **Echo – Instrument Interface Configuration Template for VBECS 09082016.gsb** file from the list.

This document is version controlled via approval signature. Do not assume that a printed, copied, or downloaded document is current. Consult the Electronic Document Control System to verify use of the current version.


6. Enter **Configuration Name** that contains 3 letter location code of the instrument (e.g. HIN for Hines VAMC), word Echo and sequence number (1 for the first instrument, 2 for the second etc.). Example **Configuration Name** for instrument configuration located at Hines would be:

**HIN\_Echo\_1**

This configuration name will be further referred in this document as *<Instrument Side Configuration>*.

7. Enter **Configuration Description** and click **Import** button. Verify that the following confirmation window displays (Figure 11).


**Figure 11: Example of Successful Configuration Import**



*If you are using newer version of the driver than the one mentioned in section 3, the Instrument Manager will warn you about the discrepancy in driver versions. Please acknowledge this warning and continue.*

8. Click **OK** and then close the **Restore Driver Configuration** window.
9. Verify that **Configuration Editor** shows the new configuration on the list (Figure 12)

**Figure 12: Example of Configuration Editor Window Showing Newly Imported Configuration**



10. Close the Configuration Editor window.

### 4.3 Verify test code mapping for instrument side configuration

1. Navigate to **Reports -> Configuration Options and Mappings**.
2. Select the **<Instrument Side Configuration Name>** from the pull down menu (Figure 13).
3. Select the **Results Test Code Mapping Tab** (Figure 13).
4. Verify the Results Test Code Mapping Report Tab matches the list in- **Appendix A: Instrument Side mapping**

**Figure 13: Example of Results Test Code Mapping Tab**

**Configuration Options and Mappings - HIN\_Echo\_1 (Echo Test)**

HIN\_Echo\_1 (Echo Test)

Enter Find Text

**Results Test Code Mapping**

**Instrument to Instrument Manager to LIS/Host (Results)**

Instrument Test Code ^	IM Test Code	Fluid	Fluid Description	Test Resulting Option	Display Name	Default Test Code	LOINC ^
Ag_C RH2-Anti-C-EW	Ag_C RH2-Anti-C-EW	Blood	Blood	00	AntiC	<input checked="" type="checkbox"/>	
Ag_c RH4-Anti-c-EW	Ag_c RH4-Anti-c-EW	Blood	Blood	00	Antic	<input checked="" type="checkbox"/>	
Ag_CcEe-Anti-c-EW	Ag_CcEe-Anti-c-EW	Blood	Blood	00	Antic	<input checked="" type="checkbox"/>	
Ag_CcEe-Anti-C-EW	Ag_CcEe-Anti-C-EW	Blood	Blood	00	AntiC	<input checked="" type="checkbox"/>	
Ag_CcEe-Anti-e-EW	Ag_CcEe-Anti-e-EW	Blood	Blood	00	Antie	<input checked="" type="checkbox"/>	
Ag_CcEe-Anti-E-EW	Ag_CcEe-Anti-E-EW	Blood	Blood	00	AntiE	<input checked="" type="checkbox"/>	
Ag_E RH3-Anti-E-EW	Ag_E RH3-Anti-E-EW	Blood	Blood	00	AntiE	<input checked="" type="checkbox"/>	
Ag_e RH5-Anti-e-EW	Ag_e RH5-Anti-e-EW	Blood	Blood	00	Antie	<input checked="" type="checkbox"/>	
Ag_Kell-Anti-Kell-EW	Ag_Kell-Anti-Kell-EW	Blood	Blood	00	AntiK	<input checked="" type="checkbox"/>	
Confirm-Anti-A-EW	Confirm-Anti-A-EW	Blood	Blood	00	AntiA	<input checked="" type="checkbox"/>	
Confirm-Anti-B-EW	Confirm-Anti-B-EW	Blood	Blood	00	AntiB	<input checked="" type="checkbox"/>	
Confirm-Anti-D-EW	Confirm-Anti-D-EW	Blood	Blood	00	AntiD	<input checked="" type="checkbox"/>	
Confirm-Confirm Interp	Confirm-Confirm Interp	Blood	Blood	00	ABORhInterp	<input checked="" type="checkbox"/>	
Confirm-Mono Ctl-EW	Confirm-Mono Ctl-EW	Blood	Blood	00	DControl	<input checked="" type="checkbox"/>	
Crossmatch-Crossmatch Interp	Crossmatch-Crossmatch Interp	Blood	Blood	00	XMIInterp	<input checked="" type="checkbox"/>	
Crossmatch-IgG XM-EW	Crossmatch-IgG XM-EW	Blood	Blood	00	AHG	<input checked="" type="checkbox"/>	
DAT-DAT Interp	DAT-DAT Interp	Blood	Blood	00	IgGInterp	<input checked="" type="checkbox"/>	
DAT-DAT Rxn-EW	DAT-DAT Rxn-EW	Blood	Blood	00	IgG	<input checked="" type="checkbox"/>	

Configuration Op... Orders Test Cod... **Results Test Cod...** Orders Fluid Cod... Results Fluid Cod... Instrument ID M... Test Load Factors Rules


**STOP** If mismatches in Test Codes names, missing or extra Test Codes are encountered, file a national CA SDM ticket to coordinate assistance with installation using the template in Appendix F. Do not proceed until the issue is resolved.



#### 4.4 Verify rules for instrument side configuration

1. Remaining in the Configuration Options and Mapping Window, select the **Rules Tab** (Figure 14).
2. Verify the Rules Tab matches the list in *Appendix C: Instrument Side Rules*.

Figure 14: Example of Rules Tab



3. Close the Configuration Options and Mappings window.


**STOP** If problems are encountered, file a national CA SDM ticket to coordinate assistance with installation using the template in Appendix F. Do not proceed until the issue is resolved.

## 4.5 Configure rules for instrument side configuration

In this section you will establish the name of the instrument associated with test results for VBECS reports. *If you are setting up multiple instruments, make sure that each has a unique name.*

1. Navigate to Configuration -> Specimen Management Configuration -> Rules Processing.
2. Select **<Instrument Side Configuration Name>** from the pull down menu (Figure 15).
3. Locate rule **SetInstrumentID**.
4. Click on the **Then** line in the rule.
5. Modify the rule by typing **<Instrument Name>** between the quotation marks in the lower box in (Figure 15).


**Figure 15: Example of Instrument Name Setup**



6. Locate rule **SetReceivingFacility**. Click on **Then** line in the rule. Modify the rule by typing **<Division Code>** between the quotation marks as shown in (Figure 16).


**i** **<Division Code>** also known as Station Number in VistA is a unique alphanumeric code that is associated with each hospital (e.g., **589** for VA Heartland West VAMC). This code must match the division code configured in **VBECS Administrator** application for a given site.

**Figure 16: Example of Division Code Setup**




- Click on the **Save Test / In Validation Rule Set** button located in the toolbar above the rules (Figure 17).


**Figure 17: Example of Save Rules Button**



8. Verify that the message in Figure 18 is received:

**Figure 18: Example of Rule Save**



 ***If problems are encountered, file a national CA SDM ticket to coordinate assistance with installation using the template in Appendix F. Do not proceed until the issue is resolved.***


9. Click on the **Save Live Rule Set** button located in the upper toolbar and click **Yes** to confirm (Figure 19).

**Figure 19: Example of Save Rules in Live Set Button**



10. Navigate to Reports->Configuration Options and Mappings and select the Rules Tab. Scroll down to ***Live – Incoming result – Before Message Queued Internally*** (Figure 20).
11. Verify that the Live Rule Set matches ***Appendix C: Instrument Side Rules*** and includes changes made to **SetReceivingFacility** (Figure 20) and **SetInstrumentID**.

**Figure 20: Example of Live Rules Set View**



**STOP** If problems are encountered, file a national CA SDM ticket to coordinate assistance with installation using the template in Appendix F. Do not proceed until the issue is resolved.

12. Close the **Configuration Options and Mappings** window.


#### 4.6 Import VBECS (HL7) side configuration

**i** Only one HL7 configuration is needed even if you use multiple Echo instruments. All Echos will share the same HL7 configuration.

**STOP** Modifying rules or test code mappings in the Instrument Manager configuration outside of this Setup Guide is not allowed and may lead to malfunction of the Automated Instrument to VBECS interface.

1. Navigate to **Configuration -> Configuration Editor** (Figure 21).
2. Click on the **Import** button.

**Figure 21: Example of Configuration Editor Window**




3. Once the **Restore Driver Configuration** window opens, click **Browse** the button and select **C:\Temp** folder (Figure 22).
4. Select **Echo – HL7 Interface Configuration Template for VBECS 09082016.gsb** file from the list. Enter **Configuration Name** that contains 3 letter location code of the instrument (e.g. **HIN** for Hines VAMC), word **VBECS** and sequence number (**1** for the first configuration, **2** for the second etc.). Example Configuration Name for VBECS side configuration located at Hines would be:

**HIN\_VBECS\_1**


This configuration name will be further referred in this document as *<HL7 Side Configuration>*.


**Figure 22: Example of Configuration Import Window**



5. Enter **Configuration Description** and click **Import** button. Verify that the confirmation window in Figure 23 displays.

**Figure 23: Example of Successful Configuration Import**




 *If you are using newer version of the driver than the one mentioned in section 3, the Instrument Manager will warn you about the discrepancy in driver versions. Please acknowledge this warning and continue.*

6. Click **OK** and close the **Restore Driver Configuration** window.
7. Verify that **Configuration Editor** shows the new configuration on the list (Figure 24).

This document is version controlled via approval signature. Do not assume that a printed, copied, or downloaded document is current. Consult the Electronic Document Control System to verify use of the current version.

**Figure 24: Example of Newly Imported HL7 Configuration**



#### **4.7 Verify test code mapping for VBECS side configuration**

5. Navigate to **Reports -> Configuration Options and Mappings**.
6. Select the **<HL7 Side Configuration Name>** from the pull down menu (Figure 25).
7. Select the **Results Test Code Mapping Tab** (Figure 25).
8. Verify the Results Test Code Mapping Report Tab matches the list in **Appendix B: HL7 (VBECS) Side Mapping**.



**Figure 25: Example of HL7 Configuration Report Window**

The screenshot shows a software window titled "IM Configuration Options and Mappings - HIN\_VBECS\_1 (VBECS side interface)". A red box highlights the window title bar. Below the title bar is a search bar labeled "Enter Find Text". The main section is titled "Results Test Code Mapping" and contains a table with the following columns: Instrument Test Code, IM Test Code, Fluid, Fluid Description, Test Resulting Option, Display Name, Default Test Code, and LOINC. The table lists various blood bank tests and their mappings. A red box highlights the "Results Test Code Mapping" tab at the bottom of the window.

Instrument Test Code	IM Test Code	Fluid	Fluid Description	Test Resulting Option	Display Name	Default Test Code	LOINC
Ag_C RH2-Anti-C-EW	AntiC	Blood	Blood	D	AntiC	<input checked="" type="checkbox"/>	
Ag_c RH4-Anti-c-EW	Antic	Blood	Blood	D	Antic	<input checked="" type="checkbox"/>	
Ag_CcEe-Anti-c-EW	Antic	Blood	Blood	D	Antic	<input checked="" type="checkbox"/>	
Ag_CcEe-Anti-C-EW	AntiC	Blood	Blood	D	AntiC	<input checked="" type="checkbox"/>	
Ag_CcEe-Anti-e-EW	Antie	Blood	Blood	D	Antie	<input checked="" type="checkbox"/>	
Ag_CcEe-Anti-E-EW	AntiE	Blood	Blood	D	AntiE	<input checked="" type="checkbox"/>	
Ag_E RH3-Anti-E-EW	AntiE	Blood	Blood	D	AntiE	<input checked="" type="checkbox"/>	
Ag_e RH5-Anti-e-EW	Antie	Blood	Blood	D	Antie	<input checked="" type="checkbox"/>	
Ag_Kell-Anti-Kell-EW	AntiK	Blood	Blood	D	AntiK	<input checked="" type="checkbox"/>	
Confirm-Anti-A-EW	AntiA	Blood	Blood	D	AntiA	<input checked="" type="checkbox"/>	
Confirm-Anti-B-EW	AntiB	Blood	Blood	D	AntiB	<input checked="" type="checkbox"/>	
Confirm-Anti-D-EW	AntiD	Blood	Blood	D	AntiD	<input checked="" type="checkbox"/>	
Confirm-Confirm Interp	ABORhInterp	Blood	Blood	D	ABORhInterp	<input checked="" type="checkbox"/>	
Confirm-Mono Ctrl-EW	DControl	Blood	Blood	D	DControl	<input checked="" type="checkbox"/>	
Crossmatch-Crossmatch Interp	XMIInterp	Blood	Blood	D	XMIInterp	<input checked="" type="checkbox"/>	
Crossmatch-IgG XMI-EW	AHG	Blood	Blood	D	AHG	<input checked="" type="checkbox"/>	
DAT-DAT Interp	IgGInterp	Blood	Blood	D	IgGInterp	<input checked="" type="checkbox"/>	
DAT-DAT Rxn-EW	IgG	Blood	Blood	D	IgG	<input checked="" type="checkbox"/>	
Group2-A1 Cells-EW	A1Cells	Blood	Blood	D	A1Cells	<input checked="" type="checkbox"/>	

9. Close **Configuration Options and Mappings** window.



*If mismatches in Test Codes names, missing or extra Test Codes are encountered, **file a national CA SDM ticket to coordinate assistance with installation using template in Appendix F.** Do not proceed until the issue is resolved.*


#### 4.8 Verify rules for VBECS (HL7) side configuration



*Modifying rules or test code mappings as imported using this Instrument Manager configuration Setup Guide may lead to malfunction of the Automated Instrument to VBECS interface.*


1. Navigate to **Configuration -> Specimen Management Configuration -> Rules Processing**.
2. On the **Rule Processing** window, select **<HL7 Side Configuration>** from the drop-down in the upper left corner.
3. Next, click on the **Quick Filter** icon located on the right side of the **Configuration Name** (Figure 26)

**Figure 26: Example of Rules Processing Toolbar**




4. On the **Quick Filter** window, check the box next to the **Outgoing Result** and click **Close** button (Figure 27).

**Figure 27: Example of Quick Filter Window**



5. Expand **Outgoing Result** section and **Before Message Sent to this Connection** sections (Figure 28).

**Figure 28: Example of Rules Window**



6. Expand all rules and verify that they match rules listed in **Appendix D: VBECS (HL7) Side Rules**.


This document is version controlled via approval signature. Do not assume that a printed, copied, or downloaded document is current. Consult the Electronic Document Control System to verify use of the current version.



***If mismatches in rules are encountered, file a national CA SDM ticket to coordinate assistance with installation using template in Appendix F. Do not proceed until the issue is resolved.***


- Click on the **Save Live Rule Set** button located in the upper toolbar and click **Yes** to confirm (Figure 29).

**Figure 29: Example of Save Rules in Live Set Button**



- Expand the **Live Rule Set** and verify that rule text matches **Appendix D: VBECS (HL7) Side Rules**. (Figure 30).

**Figure 30: Example of Live Rules Set View**



***If problems are encountered, file a national CA SDM ticket to coordinate assistance with installation using the template in Appendix F. Do not proceed until the issue is resolved.***

- Close the **Rules Setup** window.

## 5 Set Up HL7 Connection to VBECS TEST



*Modifying connection settings in the Instrument Manager outside of this Setup Guide is not allowed and may lead to malfunction of the Automated Instrument to VBECS interface.*

1. Navigate to **Configuration -> Connection Assignment**.
2. On the **Connection Assignment** window, click the **Add** button.
3. On the **Connection Properties** window, enter the **Connection Name**. Enter a name that contains **<HL7 Side Configuration>** and word **Connection**. For example:

**HIN\_VBECS\_1\_Connection**

4. Select **<HL7 Side Configuration>** as **Configuration Name**.
5. Check **Include in Specimen Management** check box.
6. Select **TCP/IP connection** in Device. (Figure 31).


**Figure 31: Example of Connection Properties Window**

The screenshot shows the 'Connection Properties - HIN\_VBECS\_1\_Connection' window. The 'Connection Name' field is highlighted with a red box and contains 'HIN\_VBECS\_1\_Connection'. The 'Configuration Name' dropdown is set to 'HIN\_VBECS\_1'. The 'Start on System Start' checkbox is checked. The 'Include in Specimen Management' checkbox in the 'Advanced Options' section is checked and highlighted with a red box. The 'TCP/IP' radio button in the 'Device' section is selected and highlighted with a red box. The 'Number of Days to Keep' section has spinners set to 3 for all categories. The 'Device Parameters' button is at the bottom left, and the 'Close' button is at the bottom right.

7. Click on **Device Parameters** button.
8. Enter **TCP/IP Address** and **TCP/IP Port Number** that matches **VBECS TEST Application IP Address and IP Port Number** configured in **VBECS TEST Administrator** application for **Auto Instrument Interface** (Figure 32). Please refer to *VistA Blood Establishment Computer Software*

(VBECS) 2.2.0 Technical Manual Security Guide for instruction on how to configure interfaces for VBECS.


**Figure 32: Example of TCP/IP Configuration Window**



The screenshot shows a window titled "IM TCP/IP Port Configuration - ProVue\_VBECS\_Connection". It contains two text input fields: "TCP / IP Address" with the placeholder text "<ip address>" and "TCP / IP Port Number" with the value "21995". Below these is a checkbox labeled "Send String When Opening Connection" which is currently unchecked. Under the checkbox is a text input field labeled "String to Send". A "Close" button is located at the bottom right of the window.

9. Click close on the **TCP/IP Port Configuration** window and click **Yes** to save.
10. Click close on the **Connection Properties** window and click **Yes** to save.
11. Verify the newly created connection shows on **Connection Assignment** window (Figure 33).

**Figure 33: Example of Newly Created Connection Figure**



The screenshot shows a window titled "IM Connection Assignment". It features a list box with the label "Connection Name". The list contains one entry, "HIN\_VBECS\_1\_Connection", which is highlighted in blue. Below the list box is a large empty rectangular area.

## 6 Set Up Instrument Connection

 Execute instructions in this section for each instrument that will be connecting to VBECS.


 *Modifying connection settings in the Instrument Manager outside of this Setup Guide is not allowed and may lead to malfunction of the Automated Instrument to VBECS interface.*

1. On the **Connection Assignment** window, click the **Add** button.
2. On the **Connection Properties** window, enter the **Connection Name**. Enter a name that contains *<Instrument Side Configuration>* and word **Connection**. For example:

**HIN\_Echo\_1\_Connection**

3. Select *<Instrument Side Configuration>* as **Configuration Name**.
4. Check **Include in Specimen Management** check box.
5. Select **TCP/IP** connection.
6. On the **Destination Lines** list check the box next to connection configured in previous section (Figure 34).

**Figure 34: Example of Connection Properties Window**



The screenshot shows the 'Connection Properties - HIN\_Echo\_1\_Connection' window. The 'Connection Name' field is 'HIN\_Echo\_1\_Connection'. The 'Configuration Name' dropdown is 'HIN\_Echo\_1'. The 'Site' and 'Location' dropdowns are empty. The 'Start on System Start' checkbox is checked. The 'Destination Line(s)' list has 'HIN\_VBECS\_1\_Connection' checked. The 'Number of Days to Keep' section has spinners for Incoming Messages, Outgoing Messages, Communications Trace, Error Messages, and Driver Data, all set to 3. The 'Advanced Options' section has 'Include in Specimen Management' checked. The 'Device' section has 'TCP/IP' selected. The 'Device Parameters ...' button is at the bottom left, and the 'Close' button is at the bottom right.

7. Click on **Device Parameters** button.
8. **TCP/IP Address** remains blank.
9. Set **TCP/IP Port Number** to *<Instrument Port>* discussed in Section **Network Connectivity Setup**.
10. Uncheck **Send String When Opening Connection** check box (Figure 35).

This document is version controlled via approval signature. Do not assume that a printed, copied, or downloaded document is current. Consult the Electronic Document Control System to verify use of the current version.

**Figure 35: Example of TCP/IP Configuration Window**

IM TCP/IP Port Configuration - Echo Milwaukee

TCP / IP Address

TCP / IP Port Number 10010

☐ Send String When Opening Connection

String to Send A

Close

11. Click close on the **TCP/IP Port Configuration** window and click **Yes** to save.
12. Click close on the **Connection Properties** window and click **Yes** to save.
13. Verify the newly created connection shows on **Connection Assignment** window (Figure 36).
14. Close the **Connection Assignment** window.

**Figure 36: Example of Newly Created Connection**

IM Connection Assignment

Connection Name


HIN\_Echo\_1\_Connection

HIN\_VBECS\_1\_Connection

## 7 Test New Connections


1. Navigate to **System -> Status**.
2. Right-click each newly created connection and choose option to **Start Selected Connections** (Figure 37).

**Figure 37: Example of Connection Status Window**



3. Verify that all newly created connections are showing **Status** of **On** after about a minute or so. (Figure 38).

**Figure 38: Example of Successful Connection Test**




**STOP** If one or more connections fail to start, file a national CA SDM ticket to coordinate assistance with installation using template in Appendix F. Do not proceed until the issue is resolved.

4. Close the **Status Display**.



## 8 Validate Instrument connectivity to VBECS TEST

Execute validation instructions from *Appendix G* to verify that Instrument is properly interfacing with VBECS.




*If one or more validation scenarios fail, file a national CA SDM ticket to coordinate assistance with installation using template in Appendix F. Do not proceed until the issue is resolved.*

## 9 Set up HL7 Connection to VBECS PROD


1. Navigate to **System -> Status**.
2. Right-click <VBECS\_Connection> and choose option to **Stop Selected Connections** (Figure 39).

**Figure 39: Example of Stop Selected Connections**



3. Wait until connection status changes to **Off**. Navigate to **Configuration -> Connection Assignment**.
4. Select <VBECS\_Connection> and click **Properties** (Figure 40).

**Figure 40: Example of Connection Assignment**




**5. On Connection Properties window click Device Parameters (Figure 41).**

**Figure 41: Example of Connection Properties**


6. Enter **TCP/IP Address** and **TCP/IP Port Number** that matches **VBECS PROD Application IP Address and IP Port Number** configured in **VBECS PROD Administrator** application for **Auto Instrument Interface** (Figure 42). Please refer to *VistA Blood Establishment Computer Software (VBECS) Technical Manual Security Guide* for instruction on how to configure interfaces for VBECS.

**Figure 42: Example of TCP/IP Configuration Window**




7. Close the **TCP/IP Port Configuration** and click **Yes** to confirm changes.
8. Close the **Connection Properties** window and click **Yes** to confirm changes.
9. Close the **Connection Assignment** window.
10. Navigate to **System -> Status**.
11. Right-click **<VBECS\_Connection>** and choose option to **Start Selected Connections** (Figure 43).


**Figure 43: Example of Connection Status Window**



12. Verify that **<VBECS\_Connection>** is showing Status of **On** after about a minute or so. (Figure 44).

**Figure 44: Example of Successful Connection Test Figure 44**



 ***If connection fails to start, file a national CA SDM ticket to coordinate assistance with installation using template in Appendix F. Do not proceed until the issue is resolved.***

13. Close the **Status Display** and Log Off the system.

## Glossary

Acronym, Term	Definition
<b>Automated Instrument</b>	Blood Bank Analyzer that performs blood testing.
<b>Division Code</b>	Also known as Station Number in Vista is the unique alphanumeric code that is associated with each hospital (e.g. 589 for VA Heartland West).
<b>Instrument Manager (IM)</b>	Software created by Data Innovations that serves as a middleware between Automated Instrument and VBECS. It translates messages containing test results sent from an instrument into HL7 messages that are then parsed into VBECS.
<b>VA</b>	Department of Veterans Affairs.
<b>VBECS</b>	VistA Blood Establishment Computer Software.
<b>VistA</b>	Veterans Health Information Systems and Technology Architecture. VistA software, developed by the VA, is used to support clinical and administrative functions at VA Medical Centers nationwide. VistA is composed of packages that undergo a verification process to ensure conformity with name spacing and other VistA standards and conventions.

## Appendices

### Appendix A: Instrument Side Mapping

Table 1- Appendix A: Instrument Side mapping

Instrument Test Code	IM Test Code	Display Name
Ag_C RH2-Anti-C-EW	Ag_C RH2-Anti-C-EW	AntiC
Ag_c RH4-Anti-c-EW	Ag_c RH4-Anti-c-EW	Antic
Ag_CcEe-Anti-C-EW	Ag_CcEe-Anti-C-EW	AntiC
Ag_CcEe-Anti-c-EW	Ag_CcEe-Anti-c-EW	Antic
Ag_CcEe-Anti-E-EW	Ag_CcEe-Anti-E-EW	AntiE
Ag_CcEe-Anti-e-EW	Ag_CcEe-Anti-e-EW	Antie
Ag_E RH3-Anti-E-EW	Ag_E RH3-Anti-E-EW	AntiE
Ag_e RH5-Anti-e-EW	Ag_e RH5-Anti-e-EW	Antie
Ag_Kell-Anti-Kell-EW	Ag_Kell-Anti-Kell-EW	AntiK
Confirm-Anti-A-EW	Confirm-Anti-A-EW	AntiA
Confirm-Anti-B-EW	Confirm-Anti-B-EW	AntiB
Confirm-Anti-D-EW	Confirm-Anti-D-EW	AntiD
Confirm-Confirm Interp	Confirm-Confirm Interp	ABORhInterp
Confirm-Mono Ctrl-EW	Confirm-Mono Ctrl-EW	DControl
Crossmatch-Crossmatch Interp	Crossmatch-Crossmatch Interp	XMInterp
Crossmatch-IgG XM-EW	Crossmatch-IgG XM-EW	AHG
DAT-DAT Interp	DAT-DAT Interp	IgGInterp
DAT-DAT Rxn-EW	DAT-DAT Rxn-EW	IgG
Group2-A1 Cells-EW	Group2-A1 Cells-EW	A1Cells
Group2-Anti-A-EW	Group2-Anti-A-EW	AntiA
Group2-Anti-B-EW	Group2-Anti-B-EW	AntiB
Group2-Anti-D1-EW	Group2-Anti-D1-EW	AntiD1
Group2-Anti-D2-EW	Group2-Anti-D2-EW	AntiD2
Group2-B Cells-EW	Group2-B Cells-EW	BCells
Group2-Group2 Interp	Group2-Group2 Interp	ABORhInterp
Group2-Mono Ctrl-EW	Group2-Mono Ctrl-EW	DControl
Group-A1 Cells-EW	Group-A1 Cells-EW	A1Cells
Group-Anti-A-EW	Group-Anti-A-EW	AntiA
Group-Anti-B-EW	Group-Anti-B-EW	AntiB
Group-Anti-D1-EW	Group-Anti-D1-EW	AntiD1
Group-Anti-D2-EW	Group-Anti-D2-EW	AntiD2
Group-B Cells-EW	Group-B Cells-EW	BCells
Group-Group Interp	Group-Group Interp	ABORhInterp
Group-Mono Ctrl-EW	Group-Mono Ctrl-EW	DControl
Screen-Screen 1-EW	Screen-Screen 1-EW	SC1
Screen-Screen 2-EW	Screen-Screen 2-EW	SC2
Screen-Screen 3-EW	Screen-Screen 3-EW	SC3
Screen-Screen Interp	Screen-Screen Interp	ABSInterp
Weak D-Anti-D-EW	Weak D-Anti-D-EW	AntiD
Weak D-Weak D Interp	Weak D-Weak D Interp	WeakDInterp

This document is version controlled via approval signature. Do not assume that a printed, copied, or downloaded document is current. Consult the Electronic Document Control System to verify use of the current version.

## **Appendix B: HL7 (VBECS) Side Mapping**

**Table 2- Appendix B: HL7 (VBECS) Side Mapping**

<b>IM Test Code</b>	<b>HL7 (VBECS) Test Code</b>
Ag_C RH2-Anti-C-EW	AntiC
Ag_c RH4-Anti-c-EW	Antic
Ag_CcEe-Anti-C-EW	AntiC
Ag_CcEe-Anti-c-EW	Antic
Ag_CcEe-Anti-E-EW	AntiE
Ag_CcEe-Anti-e-EW	Antie
Ag_E RH3-Anti-E-EW	AntiE
Ag_e RH5-Anti-e-EW	Antie
Ag_Kell-Anti-Kell-EW	AntiK
Confirm-Anti-A-EW	AntiA
Confirm-Anti-B-EW	AntiB
Confirm-Anti-D-EW	AntiD
Confirm-Confirm Interp	ABORhInterp
Confirm-Mono Ctrl-EW	DControl
Crossmatch-Crossmatch Interp	XMIInterp
Crossmatch-IgG XM-EW	AHG
DAT-DAT Interp	IgGInterp
DAT-DAT Rxn-EW	IgG
Group2-A1 Cells-EW	A1Cells
Group2-Anti-A-EW	AntiA
Group2-Anti-B-EW	AntiB
Group2-Anti-D1-EW	AntiD1
Group2-Anti-D2-EW	AntiD2
Group2-B Cells-EW	BCells
Group2-Group2 Interp	ABORhInterp
Group2-Mono Ctrl-EW	DControl
Group-A1 Cells-EW	A1Cells
Group-Anti-A-EW	AntiA
Group-Anti-B-EW	AntiB
Group-Anti-D1-EW	AntiD1
Group-Anti-D2-EW	AntiD2
Group-B Cells-EW	BCells
Group-Group Interp	ABORhInterp
Group-Mono Ctrl-EW	DControl
Screen-Screen 1-EW	SC1
Screen-Screen 2-EW	SC2
Screen-Screen 3-EW	SC3
Screen-Screen Interp	ABSInterp
Weak D-Anti-D-EW	AntiD
Weak D-Weak D Interp	WeakDInterp



## ***Appendix C: Instrument Side Rules***

### **Figure 45: Instrument Side Rules**

HIN\_Echo\_1 (Echo Test)  
ibggecoi / Immucor Galileo Echo/Arkonet / v8.00.0007

#### **--- Test / In Validation - Incoming result - Before Message Queued Internally ---**

##### **Rule # 1**

Desc - SetReceivingFacility  
If - {Always}  
Then - {Set} {Receiving Facility} = ""

##### **Rule # 2**

Desc - Set Test Name for XM  
If - {Test Code} {On Any Test} {Contains} "Crossmatch"  
Then - {Set} {Test Code Sub ID} {On That Test} = "XM"

##### **Rule # 3**

Desc - Set Test Name for DAT  
Child of Rule # 2 - Else  
If - {Test Code} {On Any Test} {Contains} "DAT"  
Then - {Set} {Test Code Sub ID} {On That Test} = "DAT"

##### **Rule # 4**

Desc - Set Test Name for Unit ABORh  
Child of Rule # 3 - Else  
If - {Test Code} {On Any Test} {Contains} "Confirm"  
Then - {Set} {Test Code Sub ID} {On That Test} = "Unit ABORh"

##### **Rule # 5**

Desc - Set Test Name for Antigen Typing  
Child of Rule # 4 - Else  
If - {Test Code} {On Any Test} {Contains} "Ag\_"  
Then - {Set} {Test Code Sub ID} {On That Test} = "Antigen Typing"

##### **Rule # 6**

Desc - Set Test Name for Weak D  
Child of Rule # 5 - Else  
If - {Test Code} {On Any Test} {Contains} "Weak D"  
Then - {Set} {Test Code Sub ID} {On That Test} = "Weak D"

##### **Rule # 7**

Desc - Set Test Name for Patient ABORh  
If - {Test Code} {On Any Test} {Contains} "Group"  
Then - {Set} {Test Code Sub ID} {On That Test} = "Patient ABORh" {AND} {Set} {Specimen User Field 15} = {Specimen User Field 15} {Concatenated With} " Group"

##### **Rule # 8**

Desc - Set Test Name for ABS  
If - {Test Code} {On Any Test} {Contains} "Screen"  
Then - {Set} {Test Code Sub ID} {On That Test} = "ABS" {AND} {Set} {Specimen User Field 15} = {Specimen User Field 15} {Concatenated With} " Screen"

Rule # 9

Desc - XMResult

If - {Result} {On Any Test} {Contains} "IgG Comp"

Then - {Set} {Result} {On That Test} = "Compatible"

Rule # 10

Desc - SetInstrumentID

If - {Always}

Then - {Set} {Instrument ID} = ""

Rule # 11


Desc - WeakDResult

If - {Result} {On Any Test} {Contains} "Positive"

Then - {Set} {Result} {On That Test} = "Positive"

## Appendix D: VBECS (HL7) Side Rules

Figure 46: Rules: HL7 Side Rules



## Appendix E: Echo Testing Templates

**Table 3: Echo Testing Templates**

Template Name	Selected Fields	VBECS Test	Test Component
Group	Group-Anti-A-EW Group-Anti-B-EW Group-A1 Cells-EW Group-B Cells-EW Group-Anti-D1-EW Group-Anti-D2-EW Group-Mono Ctrl-EW Group-Group Interp	Patient ABO/Rh	Anti-A Anti-B A1 Cells B Cells Anti-D  D Control ABO Interp Rh Interp
Group2	Group2-Anti-A-EW Group2-Anti-B-EW Group2-A1 Cells-EW Group2-B Cells-EW Group2-Anti-D1-EW Group2-Anti-D2-EW Group2-Mono Ctrl-EW Group2-Group Interp	Patient ABO/Rh	Anti-A Anti-B A1 Cells B Cells Anti-D  D Control ABO Interp Rh Interp
Screen	Screen-Screen 1-EW Screen-Screen 2-EW Screen-Screen 3-EW Screen-Screen Interp	Antibody Screen	Screening Cells1 Screening Cells2 Screening Cells 3 ABS Interp
Group Screen	Group-Anti-A-EW Group-Anti-B-EW Group-A1 Cells-EW Group-B Cells-EW Group-Anti-D1-EW Group-Anti-D2-EW Group-Mono Ctrl-EW Group-Group Interp Screen-Screen 1-EW Screen-Screen 2-EW Screen-Screen 3-EW Screen-Screen Interp	Type and Screen	Anti-A Anti-B A1 Cells B Cells Anti-D  D Control ABO Interp Rh Interp Screening Cells1 Screening Cells2 Screening Cells 3 ABS Interp
DAT	DAT-DAT Rxn-EW DAT-DAT Interp	DAT IgG	IgG DAT IgG Interp
Crossmatch	Crossmatch-IgG XM-EW Crossmatch-Crossmatch Interp	Crossmatch	AHG XM Interp
Confirm	Confirm-Anti-A-EW Confirm-Anti-B-EW Confirm-Anti-D-EW Confirm-Mono Ctrl-EW Confirm-Confirm Interp	Unit ABO/Rh Confirmation	Anti-A Anti-B Anti-D D Control ABO Interp Rh Interp
Weak D	Weak D-Anti-D-EW Weak D-Weak D Interp	Patient and Unit Weak D	Anti-D Weak D Interp
Ag_CcEe	Ag_CcEe-Anti-C-EW Ag_CcEe-Anti-E-EW Ag_CcEe-Anti-c-EW Ag_CcEe-Anti-e-EW	Patient and Unit Antigen Typing	Anti-C Anti-E Anti-c Anti-e
Ag_C RH2	Ag_C RH2-Anti-C-EW	Patient and Unit Antigen Typing	Anti-C
Ag_c RH4	Ag_c RH4-Anti-c-EW	Patient and Unit Antigen Typing	Anti-c
Ag_E RH3	Ag_E RH3-Anti-E-EW	Patient and Unit Antigen Typing	Anti-E

This document is version controlled via approval signature. Do not assume that a printed, copied, or downloaded document is current. Consult the Electronic Document Control System to verify use of the current version.

Template Name	Selected Fields	VBECS Test	Test Component
Ag_e RH5	Ag_e RH5-Anti-e-EW	Patient and Unit Antigen Typing	Anti-e
Ag_Kell	Ag_Kell-Anti-Kell-EW	Patient and Unit Antigen Typing	Anti-K


## ***Appendix F: CA SDM Support Ticket Template***

Please use the following to complete your ticket:

- **Incident area:** NTL.APP.HealtheVet VistA.VBECS 2\_0
- **Summary:** VBECS 2.2.0 Automated Instrument configuration assistance request
- **Description:**
  - Enter contact information for the main blood bank and the name of the responsible individual(s) for all communications.
  - Enter preferred date and time.
- **Property Questions:** Respond NO to the required questions.
- **Group:** automatically fills in correctly do not change it from. *NTL Alert Blood Bank & VBECS*
- **Add additional information as required.**

## Appendix G: Validation Planning and Example Test Scenarios

The following is a flowchart to help assess any one change and plan accordingly.





*These are examples of test scenarios. Each site is responsible for evaluating changes for their intended use, assess risk, and for establishing additional validation test scenarios.*

*All test scenarios have a suggested user role, this may require adjustment to align with the patient or unit data selected to execute the scenario. Process any overrides as well, based on patient or unit selection.*

*\*The Expected Outcome numbering corresponds to the Step number where the change verification appears.*

## Test Group One: AI interface disabled

**Test Objective:** Demonstrate that the system will reject test results sent from automated instrument if the Automated Instrument interface is disabled.

**Option:** VBECS Administrator, Automated Instrument Interface

**At least one instrument available in any division and configured for connection with VBECS via Instrument Manager.**

**Note:** This may be executed at only one division if VBECS is used in a multidivisional configuration as the interface is enabled/disabled for all or none.

Data	<p>Before beginning, verify current configuration, activity, and status of the existing interfaces, VistALink, and CPRS in your test account.          Make sure that the VBECS-OERR HL7 link in VistA is NOT shutdown.</p> <ol style="list-style-type: none"> <li>1. Create RBC and TAS orders in CPRS and accept them in VBECS.</li> </ol>
User	<p>VBECS Administrator access is required to configure the AUTOMATED TESTING interface.          No specific user role is required to process TAS and RBC orders in VBECS.</p>
Steps	<ol style="list-style-type: none"> <li>1. Log into VBECS Administrator:</li> <li>2. Disable the AUTOMATED TESTING interface using Configure Interfaces option.</li> <li>3. A message appears (You are about to disable Auto Instrument Interface. It will cause VBECS to stop sending and receiving messages via that interface. Continue?), Click Yes.</li> <li>4. Close the window as all fields are disabled.</li> <li>5. Perform TAS test on the instrument and send results of testing to VBECS</li> </ol>
Expected Outcome	<ol style="list-style-type: none"> <li>5. Verify that after sending test results from an instrument fails it is possible to complete TAS test using manual testing grids and subsequently issue and transfuse blood unit to a patient.</li> </ol>
Reports:	<ol style="list-style-type: none"> <li>6. Review the Audit Trail report for changes to the interface.</li> </ol>



## Test Group Two: Verify AI individual test(s)

**Test Objective:** Demonstrate that the system will allow the selected user role to perform normal workflow activities using your local blood bank testing instrument.

**Note:** Validate all tests that are performed at your site using an automated instrument.



Prior to initiating any testing of the automated instrument interface, see the VistA Blood Establishment Computer Software (VBECS) 2.2.0 Technical Manual-Security Guide for instructions for configuring your local instrument and Data Innovations Instrument Manager.



Configure your Test UID to clearly differentiate from a Production UID during your validation, for example, production is 1234567890, with 123 being your consistent prefix, change 123 or 12 to TS or TST or 999. Please change the Numeric Identifier for the Blood Bank Accession Area in your site Test Account.

**Enable Automated Instrument Interface in VBECS Admin if it was disabled while performing testing in Test Group One.**

Verify your VBECS processes with the Automated Instrument interface.

Your local test plan will demonstrate that the system will perform normal daily work per your local policies, procedures and local validation plan that may include:

Blood Units: Automated Instrument or via the short cut key 

- ABO-Rh confirmation
- Unit Antigen Typing or Repeat

Patients: Automated Instrument or via the short cut key 

- Perform a Type and Screen test (ABS and ABS with Auto-control only)
- Perform crossmatch tests: Serologic (*Selected in Blood Units: Select Units before testing.\**)
- Patient Tests:
  - ABO/Rh or Repeat
  - Antibody Screen or Repeat
  - Direct Antiglobulin Test or Repeat
  - Antigen Typing or Repeat
- Overrides, (ABO/Rh Discrepancy, Crossmatch incompatible: Give with Medical Director Approval)
- Report data from these actions is available for retrieval.
  - Testing Worklist Report
  - Patient History Report
  - Unit History Report
  - Exception Report (ABO/Rh Discrepancy, Give with Medical Director Approval)

\*Remember to click NO to proceeding to the serologic crossmatch when selecting units for automated instrument testing.

## ***Test Group 2 Scenario 1: Verify AI TAS test***

**Note:** the Type and Screen (TAS) is a combination of two tests and may be saved as individual tests without completing the TAS as a whole.

<b>Data</b>	VBECS: Select a patient. CPRS: Place a Type and Screen (TAS) order for the patient LAB: Accession the order VBECS: Accept the order. (Orders, Accept Orders)
<b>User</b>	No specific user role is required to process TAS order in VBECS.
<b>Steps</b>	<ol style="list-style-type: none"> <li>1. User checks the Patient testing list to make sure the order is accepted in VBECS and appears on the Pending Task List (PTL). (Patient, Patient Testing, Diagnostic Tests).</li> <li>2. Close the PTL.</li> <li>3. Process the specimen on the instrument using the recommended TAS template for that instrument. Complete all work needed to transmit the information to VBECS.</li> <li>4. Select Patients, Automated Instrument to review TAS results.</li> <li>5. Select the specimen UID, scanning the UID is preferred.</li> <li>6. Review TAS test results.</li> <li>7. Accept only the ABS or ABO/Rh test. Close the window.</li> <li>8. Open the PTL.</li> <li>9. Try to select TAS on PTL and complete testing on it manually.</li> <li>10. Open the Automated instrument window and accept the second part of TAS.</li> <li>11. Close the automated instrument window.</li> <li>12. Check Reports</li> </ol>
<b>Expected Outcome</b>	<ol style="list-style-type: none"> <li>5. Verify that the specimen UID is selectable by scanning, entry or patient selection</li> <li>6. Verify that the correct test results appear on the Automated Instrument review list.</li> <li>8. Verify that the TAS appears on the PTL with a status of "Instrument Results Pending Review".</li> <li>9. Verify that system prevents user from completing TAS since they are still pending results from an instrument for it.</li> <li>12. Verify that the results and comments appear as expected on the reports: <ul style="list-style-type: none"> <li>• Testing Worklist Report</li> <li>• Patient History Report</li> <li>• Exception Report</li> </ul> </li> </ol>

<b><i>Test Group 2 Scenario 2: Verify AI serologic crossmatch test</i></b>	
<b>Data</b>	<p>VBECS: Select a patient.  CPRS: Place a Type and Screen (TAS) order and Red Blood Cell order for the patient  LAB: Accession the orders</p> <p>VBECS: Accept the order. (Orders, Accept Orders)</p> <p>Process the TAS to completion.</p> <p>Select a blood unit for a selected patient*:</p> <ul style="list-style-type: none"> <li>• Previously entered into the division's inventory (Blood Units, Incoming Shipment)</li> <li>• ABO compatible</li> <li>• May be available or selected for another patient (available, selected, crossmatched) status.</li> <li>• May or may not trigger selection overrides</li> </ul> <p>*Remember to click NO to proceeding to the serologic crossmatch when selecting units for automated instrument testing.</p>
<b>User</b>	No specific user role is required to process crossmatch test in VBECS.
<b>Steps</b>	<ol style="list-style-type: none"> <li>1. Process the component unit's specimen on the instrument using the recommended template for that instrument. Complete all work needed to transmit the information to VBECS.</li> <li>2. Select Patients, Automated Instrument to review crossmatch results.</li> <li>3. Select the specimen UID, scanning the UID is preferred.</li> <li>4. Review results, select compatibility and accept crossmatch result.</li> <li>5. Print or do not print tag as desired.</li> <li>6. Accept the test and close the automated instrument window.</li> <li>7. Check Reports.</li> </ol>
<b>Expected Outcome</b>	<ol style="list-style-type: none"> <li>4. Verify that crossmatch results sent from an instrument show correctly on the screen.</li> <li>7. Verify that the results and comments appear as expected on the reports: <ul style="list-style-type: none"> <li>• Testing Worklist Report</li> <li>• Patient History Report (interpretations only).</li> <li>• Exception Report</li> </ul> </li> </ol>

<b>Test Group 2 Scenario 3: Verify AI patient diagnostic tests</b> (ABO/Rh, Antibody Screen, Direct Antiglobulin Test, Patient Antigen Typing, and the reflex test)	
<b>Data</b>	VBECS: Select a patient. CPRS: Place a diagnostic test order for the patient LAB: Accession the order VBECS: Accept the order. (Orders, Accept Orders)
<b>User</b>	No specific user role is required to process diagnostic tests in VBECS.
<b>Steps</b>	<ol style="list-style-type: none"> <li>1. User checks the Patient testing list to make sure the order is accepted in VBECS and appears on the Pending Task List (PTL). (Patients, Patient Testing, Diagnostic Tests).</li> <li>2. Close the PTL.</li> <li>3. Process the specimen on the instrument using the recommended testing template for that instrument. Complete all work needed to transmit the information to VBECS.</li> <li>4. Select Patients, Automated Instrument to review test results.</li> <li>5. Select the specimen UID, scanning the UID is preferred.</li> <li>6. Review and accept test results</li> <li>7. Close the automated instrument window.</li> <li>8. Check reports.</li> </ol>
<b>Expected Outcome</b>	<ol style="list-style-type: none"> <li>5. Verify that the specimen UID is selectable by scanning, entry or patient selection.</li> <li>6. Verify that patient test results sent from an instrument show correctly on the screen.</li> <li>8. Verify that the results and comments appear as expected on the reports: <ul style="list-style-type: none"> <li>• Testing Worklist Report</li> <li>• Patient History Report</li> <li>• Exception Report</li> </ul> </li> </ol>

<b><i>Test Group 2 Scenario 4: Verify AI blood unit tests</i></b>	
<b>Data</b>	<p>VBECS: Select a blood unit previously entered into the division's inventory (Blood Units, Incoming Shipment).</p> <p>For ABO/Rh Confirmation testing on the instrument, unit should be in a Limited status.</p> <p>For Unit Antigen Typing, the unit may or may not have been confirmed.</p>
<b>User</b>	No specific user role is required to process blood unit tests in VBECS.
<b>Steps</b>	<ol style="list-style-type: none"> <li>1. Process the component unit's specimen on the instrument using the recommended template for that instrument. Complete all work needed to transmit the information to VBECS.</li> <li>2. Select Blood Units, Automated Instrument to review test results.</li> <li>3. Select the blood component unit's Donor Identification Number (DIN), scanning the DIN is preferred.</li> <li>4. Select the product code (if there are multiple blood units with the same product code)</li> <li>5. Review the transmitted blood unit test.</li> <li>6. Accept the test and close the automated instrument window.</li> <li>7. Check Reports</li> </ol>
<b>Expected Outcome</b>	<ol style="list-style-type: none"> <li>5. Verify that blood unit test results show correctly on the screen.</li> <li>7. Verify that the results and comments appear as expected on the reports: <ul style="list-style-type: none"> <li>• Testing Worklist Report</li> <li>• Unit History Report (interpretations only).</li> <li>• Exception Report</li> </ul> </li> </ol>