

# **Technical Recommendations**

### **Tool Reference**

RST Reference Number: RST24CV14.01

Date of Publication: 02/07/2024

Recommended Citation: U.S. Food and Drug Administration. (2024). *Mock Circulatory Loop to Generate Variable Adult Heart Conditions for Evaluating Mechanical Circulatory Support Devices* (RST24CV14.01). <a href="https://cdrh-rst.fda.gov/mock-circulatory-loop-generate-variable-adult-heart-conditions-evaluating-mechanical-circulatory">https://cdrh-rst.fda.gov/mock-circulatory-loop-generate-variable-adult-heart-conditions-evaluating-mechanical-circulatory</a>

### For more information

Catalog of Regulatory Science Tools to Help Assess New Medical Devices



# **Disclaimer**

### **About the Catalog of Regulatory Science Tools**

The enclosed tool is part of the Catalog of Regulatory Science Tools, which provides a peer-reviewed resource for stakeholders to use where standards and qualified Medical Device Development Tools (MDDTs) do not yet exist. These tools do not replace FDA-recognized standards or MDDTs. This catalog collates a variety of regulatory science tools that the FDA's Center for Devices and Radiological Health's (CDRH) Office of Science and Engineering Labs (OSEL) developed. These tools use the most innovative science to support medical device development and patient access to safe and effective medical devices. If you are considering using a tool from this catalog in your marketing submissions, note that these tools have not been qualified as Medical Device Development Tools and the FDA has not evaluated the suitability of these tools within any specific context of use. You may request feedback or meetings for medical device submissions as part of the Q-Submission Program.

For more information about the Catalog of Regulatory Science Tools, email RST\_CDRH@fda.hhs.gov.



## **Technical Recommendations**

The following recommendations are complementary to the ISO 14708-5:2020 specifications for simulating well-defined target test conditions in a reproducible manner using mock circulatory loops (MCLs):

- 1. Describing the test equipment [refer to **Section 2.1** in <u>D'Souza et al., 2024</u>]
- 2. Defining target clinical use test conditions [refer to Section 2.2 in D'Souza et al., 2024]
- 3. Specifying output quantities of interest [refer to Section 2.3 in D'Souza et al., 2024]
- 4. Conducting MCL simulations of the target test conditions [refer to **Section 3** in <u>D'Souza et al., 2024</u>; and MCL User Manual]
- 5. Characterizing the repeatability of the MCL [refer to MCL Repeatability Testing]