



«O`zbekiston ilmiy – sinov va sifat nazorati markazi» DM
qoshidagi Sinov laboratoriyalar majmuasi

Elektromagnit
moslashuvchanlik
laboratoriyasi

Po`chta manzili:

111404, Toshkent viloyati,
Piskent tumani,
Lola-ariq MFY,
O`zbekiston ko'chasi,
174-uy.

«Цифровой мультиметр»

Model: *HMC 8012*

Zavod raqami: *№ 106900*

Ishlab chiqaruvchi: *Chexiya.*

O'lchov chegarasi, V : *om 0,4 do 750 V*

Xatolik darajasi: *0,015 %*

Ta'minot kuchlanishi: *220/230 $V \pm 10$ %.*

Ta'minot manbai chastotasi: *50/60 $Hz \pm 0.5$*

Ish harorat, namlik oralig'i va atmosfera bosimi: *(20 ± 5) $^{\circ}C$, (30-80) %, (84-106.7) kPa*

