



«O`zbekiston ilmiy – sinov va sifat nazorati markazi» DM
qoshidagi Sinov laboratoriyalar majmuasi

Elektromagnit
moslashuvchanlik
laboratoriyasi

Po`chta manzili:

111404, Toshkent viloyati,
Piskent tumani,
Lola-ariq MFY,
O`zbekiston ko'chasi,
174-uy.

«Цифровой мультиметр»

Model: *HMC 8012*

Zavod raqami: № 106891

Ishlab chiqaruvchi: *Chexiya.*

O'lchov chegarasi, V: от 0,4 до 750 V

Xatolik darajasi: 0,015 %

Ta'minot kuchlanishi: 220/230 V ± 10 %.

Ta'minot manbai chastotasi: 50/60 Hz ± 0.5

Ish harorat, namlik oralig'I va atmosfera bosimi: (20 \pm 5) °C, (30-80) %, (84-106.7) kPa

