



«O`zbekiston ilmiy – sinov va sifat nazorati markazi» DM
qoshidagi Sinov labaratoriylar majmuasi

Elektromagnit
moslashuvchanlik
laboratoriyasi

Po`chta manzili:

111404, Toshkent viloyati,
Piskent tumani,
Lola-ariq MFY,
O`zbekiston ko‘chasi,
174-uy.

«Поглощающие клещи»

Model: MDS21

Zavod raqami: № 100978

Ishlab chiqaruvchi: Germaniya.

O`lchov chegarasi, Hz: (30 MHz - 6 GHz)

Xatolik darajasi: $\pm 5dB$

Ish harorat, namlik oralig'i va atmosfera bosimi: $(20 \pm 5) ^\circ C$, (30-80) %, (84-106.7) kPa

