



«O`zbekiston ilmiy – sinov va sifat nazorati markazi» DM  
qoshidagi Sinov laboratoriyalar majmuasi

Elektromagnit  
moslashuvchanlik  
laboratoriyasi

Po`chta manzili:

111404, Toshkent viloyati,  
Piskent tumani,  
Lola-ariq MFY,  
O`zbekiston ko`chasi,  
174-uy.

## «Термогигрометр»



**Model:** TFA 30.5002

**Zavod raqami:** № H371

**Ishlab chiqaruvchi:** Germaniya.

**O'lchov chegarasi,** (10 - 99) % , -10 °C do + 60 °C

**Xatolik darajasi:** ± 5 %, ±1°C

**Ish harorat, namlik oralig'I va atmosfera bosimi:** (20 ±5) °C, (30-80) %, (84-106.7) kPa